



# PicoScope® 9300 Series

## PC Sampling Oscilloscopes

Programmer's Guide



# Contents

1 PicoScope 9300 API Reference .....	1
1 PicoScope9300 COM Server .....	1
2 ExecCommand Method .....	1
3 COMRC Object .....	1
2 Commands Syntax .....	2
1 Command and Query Structure .....	2
1 Overview .....	2
2 Messages .....	2
3 Commands .....	3
4 Queries .....	3
5 Headers .....	3
2 Command Entry .....	4
1 Rules .....	4
2 Concatenation .....	4
3 Command Classification .....	6
1 Execution-type commands .....	6
2 On/off-type commands .....	6
3 On/off-group-type commands .....	6
4 Selector-type commands .....	8
5 Integer-type commands .....	8
6 Float-type commands .....	9
7 Data-type commands .....	9
4 Full list of commands .....	10
1 Header commands .....	10
2 GUI commands .....	10
3 System commands .....	11
4 Channels commands .....	13
5 Timebase commands .....	17
6 Trigger commands .....	20
1 Trigger Source .....	20
2 External Direct Trigger Commands .....	20
3 External Prescale Trigger Commands .....	20
4 Ch1/Ch2 Direct Trigger Commands .....	21
5 Trigger Period for Internal Clock Sources .....	22
6 Trigger Mode .....	22
7 Trigger Holdoff Commands .....	22
8 Attenuator Commands .....	23
9 Pattern Sync Trigger Commands .....	23
7 Acquisition commands .....	25
1 Sampling Mode .....	25

<b>2 Channel Acquisition Commands .....</b>	<b>25</b>
<b>3 Termination of the Acquisition .....</b>	<b>26</b>
<b>4 Number of Waveforms .....</b>	<b>26</b>
<b>5 Action when Number of Waveforms reached .....</b>	<b>26</b>
<b>6 File Name .....</b>	<b>27</b>
<b>7 Stored Files Format .....</b>	<b>27</b>
<b>8 Display commands .....</b>	<b>28</b>
<b>9 Save/Recall commands .....</b>	<b>31</b>
<b>1 Work with Memo Zones (M1, M2, M3, M4) .....</b>	<b>31</b>
<b>2 Memory Scaling .....</b>	<b>31</b>
<b>3 Work with Disk .....</b>	<b>33</b>
<b>4 Work with Setups .....</b>	<b>35</b>
<b>10 Markers commands .....</b>	<b>37</b>
<b>11 Measure commands .....</b>	<b>38</b>
<b>1 Common Measures commands .....</b>	<b>38</b>
<b>2 Statistic Commands .....</b>	<b>39</b>
<b>3 Define parameter Commands .....</b>	<b>39</b>
<b>4 List of Measurements .....</b>	<b>42</b>
<b>5 Inter-signal Measurements .....</b>	<b>43</b>
<b>6 Measurements of Spectrum Signals .....</b>	<b>43</b>
<b>7 Delete all Measurements for all Sources .....</b>	<b>45</b>
<b>8 Getting Measurement Results .....</b>	<b>45</b>
<b>12 Limit Tests commands .....</b>	<b>46</b>
<b>1 Limit Test On/Off .....</b>	<b>46</b>
<b>2 Limit Test Termination Commands .....</b>	<b>46</b>
<b>3 Limit Test Action Commands .....</b>	<b>46</b>
<b>4 Parameter Definition Commands .....</b>	<b>47</b>
<b>13 Mathematics commands .....</b>	<b>50</b>
<b>1 Enable Mathematical Function .....</b>	<b>50</b>
<b>2 DisplayMathematical Function .....</b>	<b>50</b>
<b>3 Function Category .....</b>	<b>50</b>
<b>4 Function Operators .....</b>	<b>50</b>
<b>5 Function Operands .....</b>	<b>52</b>
<b>6 Additional Parameters for Arithmetic Functions .....</b>	<b>53</b>
<b>7 Additional Parameters for Algebraic Functions .....</b>	<b>54</b>
<b>8 Additional Parameters for Trigonometric Functions .....</b>	<b>54</b>
<b>9 Additional Parameters for FFT Functions .....</b>	<b>54</b>
<b>10 Additional Parameters for Bit Functions .....</b>	<b>55</b>
<b>11 Additional Parameters for Miscellaneous Functions .....</b>	<b>56</b>
<b>12 Function Scaling .....</b>	<b>57</b>
<b>14 Histogram commands .....</b>	<b>59</b>
<b>1 General Histogram Commands .....</b>	<b>59</b>
<b>2 Histogram Completion Commands .....</b>	<b>59</b>

3 Histogram Window Commands .....	60
4 Histogram Calculation Commands .....	61
5 Histogram Scale Commands .....	61
6 Histogram Result Commands .....	63
15 Eye Diagram commands .....	64
1 General Eye Commands .....	64
2 Eye Measurements Commands .....	64
3 Define Parameters Commands .....	66
4 Eye Calculation Commands .....	67
5 Getting Eye Measurement Results .....	68
16 Mask Test commands .....	69
1 Common Mask Test Commands .....	69
2 Mask Creating .....	69
3 Standard Mask Test Commands .....	70
4 Automask Commands .....	72
5 Mask Test Termination .....	73
6 Mask Test Actions .....	74
7 User Mask .....	74
8 Getting Mask Test Results .....	75
17 Autocalibration commands .....	77
1 Single-shot Autocalibration .....	77
2 Periodic Autocalibration .....	78
3 Balancing the channels manually .....	79
18 Waveforms commands .....	80
19 Optical commands (PicoScope 9321 only) .....	82
1 Common optical commands .....	82
2 User-defined calibration points commands .....	83
3 Dark Level calibration commands .....	83
4 Calibration values saving commands .....	84
20 TDR/TDT Commands .....	85
1 Stimuli source command (PicoScope 9311 and 9312 only) .....	85
2 Common Internal Stimuli commands (PicoScope 9311/9312 only) .....	85
3 PicoScope 9311 Internal Stimuli commands .....	86
4 PicoScope 9312 Internal Stimuli commands .....	86
5 Reflectometer Setup commands .....	87
21 Instrument Info Commands .....	91
5 Revision history .....	94
Index .....	97



# 1 PicoScope 9300 API Reference

PicoScope 9000 provides an API for any third-party application or library to control the oscilloscope and collect signals. The API is *COM-based* and is provided by the PicoScope 9300 GUI application.

## 1.1 PicoScope9300 COM Server

The COM server implementing the API is called *PicoScope9300* and is implemented by the PicoScope 9300 GUI application (*PicoScope9300.exe*). It is registered in the system during the setup process, and can be explicitly unregistered and registered again by executing *PicoScope9000.exe* with the /UnregServer or /RegServer switches.

## 1.2 ExecCommand Method

The *COMRC* object contains only one method, *ExecCommand*. This method has one argument, a text string with a command or query. The method returns:

- *NULL* (*Nothing* in Visual Basic) if a command without query has been successfully executed
- The text string *ERROR* if the command was invalid
- Another text string with query results if either the command was a query or a command with a query

The syntax of the commands and queries and the full list of commands are described in the following pages.

## 1.3 COMRC Object

To implement the API the server exposes only one object, *COMRC*. This object supports automation, so it can be used by high-level languages like JavaScript (HTML pages) or VBA (Microsoft Word). Additionally, low-level languages like C are also supported. The string defining the system-wide name of the object and used for object creation is *PicoScope9000.COMRC*.

## 2 Commands Syntax

### 2.1 Command and Query Structure

#### 2.1.1 Overview

The PicoScope 9300 commands consist of set commands and query commands (usually called commands and queries).

- Commands modify instrument settings or tell the instrument to perform a specific action.
- Queries cause the instrument to return data and information about its status.

Most commands have both a set form and a query form. The query form of the command differs from the set form by the addition of a question mark at the end. For example, the set command

`ACQuire:Ch1:MODE`

has a query form

`ACQuire:Ch1:MODE?`

Not all commands have both a set and a query form. Some may have set only and some have query only.

#### 2.1.2 Messages

A command message is a command or query name followed by any information the instrument needs to execute the command or query. Command messages may contain five element types, as defined in the following table.

Symbol	Meaning
<code>&lt;Header&gt;</code>	This is the basic command name. If the header ends with a question mark, the command is a query. If the command is concatenated with other commands, the header must begin with a colon (:).
<code>&lt;Mnemonic&gt;</code>	This is the header of the sub-function. Some command headers have only one mnemonic. If a command header has multiple mnemonics, a colon (:) character always separates items from one another.
<code>&lt;Argument&gt;</code>	This is a quantity, quality, restriction or limit associated with the header. Some commands have no arguments while others have multiple arguments. A space separates arguments from the header. A comma separates arguments from one another.
<code>&lt;Comma&gt;</code>	A single comma is used between the arguments of multiple-argument commands. Optionally, there may be white space characters before and after the comma.
<code>&lt;Space&gt;</code>	A white space character is used between a command header and its argument. Optionally, a white space may consist of multiple white space characters.

#### Command message elements

### 2.1.3 Commands

Commands cause the instrument to perform a specific function or change one of its settings. Commands have this structure:

```
[ :]<Header>[ <Space><Argument>[ <Comma><Argument>] . . . ]
```

A command header consists of one or more mnemonics arranged in a hierarchy or tree structure. The first mnemonic is the base or root of the tree and each subsequent mnemonic is a level or branch off the previous one. Commands at a higher level in the tree may affect those at a lower level. The leading colon (:) always returns you to the base of the command tree.

### 2.1.4 Queries

Queries cause the instrument to return information about its status or settings. Queries have the structure:

- [ :]<Header>?
- [ :]<Header>?[ <Space><Argument>[ <Comma><Argument>] . . . ]

You can specify a query command at any level within the command tree unless otherwise noted. These branch queries return information about all the mnemonics below the specified branch or level. For example

```
Histogram:STATistics:STDdev?
```

returns the standard deviation of the histogram, whereas

```
Histogram:STATistics?
```

returns all the histogram statistics, and

```
Histogram?
```

returns all the histogram parameters.

### 2.1.5 Headers

You can control whether the instrument returns headers as part of the query response. Use the HEADER command to control this feature. If header is on, the query response returns command headers and formats itself as a valid set command. When the header is off, the response includes only the values. This may make it easier to parse and extract the information from the response. The table below shows the difference in responses.

Query	Header Off	Header On
Ch1:Scale?	200 mV/div	CH1:SCALE 200 mV/div
Acq:Ch1:RecLen?	512	ACQ:CH1:RECLEN 512

#### Comparison of Header Off and Header On responses

## 2.2 Command Entry

### 2.2.1 Rules

The following rules apply when entering commands:

- A mnemonic can be followed by any letters for easier understanding of the program's text. For example, these commands are all equivalent:

Ch1:ATTEN:DIMENS Volt

Ch1:ATTENUATOR:DIMENSION Volt

Ch1:ATTENblabla:DIMENSblabla Volt

However, arguments must not be followed by additional characters.

- You can enter commands in upper or lower case.
- You can precede any command with white space characters. White space characters include any combination of the ASCII control characters 00 to 09, and 0B to 20 hexadecimal (0 to 9, and 11 to 32 decimal).
- The instrument will ignore commands consisting of any combination of white space characters and line feeds.

### 2.2.2 Concatenation

You can concatenate any combination of set commands and queries by using a semicolon (;). The instrument executes concatenated commands in the order received. The following rules apply when concatenating commands and queries:

- You can separate completely different headers with a semicolon (;), and by adding a leading colon (:) at the beginning of all commands except the first one. For example

TRIGger:MODE FREE  
ACQuire:NUMAVg 10

can be concatenated into the single command

TRIGger:MODE FREE;:ACQuire:NUMAVg 10

- If concatenated commands have headers that differ by only the last mnemonic, you can abbreviate the second command and eliminate the leading colon. For example, you can concatenate the commands

ACQuire:Ch1:MODE ENVMINMAX  
ACQuire:Ch1:NAVg 10

into a single command

ACQuire:Ch1:MODE ENVMINMAX; NAVG 10

The longer version also works equally well:

ACQuire:CH1:MODE ENVMINMAX;:ACQuire:Ch1:NAVg 10

- Set commands and queries may be concatenated in the same message. For example:

```
ACQuire:CH1:MODE AVGSTAB;NAVG?
```

This is a valid message that sets the acquisition mode to Stable Averaging. The message then queries the number of acquisitions for averaging. Concatenated commands and queries are executed in the order received.

- Here are some invalid concatenations:

**DISPLAY:STYLE DOTS;ACQuire:NAVG 10**  
(a colon is needed before ACQuire)

**DISPLAY:STYLE DOTS;:FORMAT YT**  
(there is an extra colon before FORMAT. Use DISPLAY:STYLE DOTS;FORMAT YT instead.)

**Acq:Ch1:Mode Sample;Ch1:RecLen 1024**  
(The levels of these mnemonics are different. Either remove the second use of Ch1: or place :Acq: in front of Ch1:.)

## 3 Command Classification

Most commands belong to one of a few types. For example, execution-type commands tell the instrument to perform a specific action, selector-type commands modify a specific instrument setting to the one of few fixed values, and so on. All commands of a given type have similar behavior.

### 3.1 Execution-type commands

Execution-type commands tell the instrument to perform a specific action. For example:

```
*Run
*ClrDispL
```

There are no arguments for these commands.

All execution-type commands have a set form only, with no *query* form.

### 3.2 On/off-type commands

On/off type commands tell the instrument to turn on or turn off a specific function. For example:

```
Header Off
Ch1:Display 0
```

There are four fixed arguments possible in these commands: On, Off, 0, 1. Arguments On and 1 are equivalent and turn on the corresponding function. Arguments Off and 0 are also equivalent and turn off the corresponding function.

All on/off type commands have a query form, which will return one of two fixed values: ON or OFF. It is also possible to use the query form with an argument. For example:

```
Ch1:Display? 0
```

This command turns off the graphic of Channel 1 and returns OFF.

### 3.3 On/off-group-type commands

Some functions of the instrument have items that may be set on or off independently. It is also possible for the items to be either all on or all off. An example of this type of command is:

```
Meas:Ch1:XParam
```

This command has a set of parameters for automatic X-axis measurements for Ch1. It is possible to select up to 10 parameters from a list of 18:

```
Period, Freq, PosWidth, NegWidth, Rise, Fall, PosDuty,
NegDuty, PosCross, NegCross, BurstWidth, Cycles,
TimeOfMax, TimeOfMin, PosJitterPp, PosJitterRMS,
NegJitterPp, NegJitterRMS
```

There are between 2 and 64 custom items in the on/off-group-type commands. The full set of items for each command is specified in the list of commands.

The on/off-group-type commands can be used in several modes. Every such command can be used in every mode.

### **Single-item mode**

Single-item mode is used to control one item of a command without changing its other items. In this case the item's mnemonic is added to the end of the command after a colon (:). This must be followed by a space character and then one of the following arguments: On, Off, 0, 1. For example, this command turns on a frequency measurement for Channel 1:

```
Meas:Ch1:XParam:Freq 1
```

Single-item mode has a query form similar to the On/off commands. So the query

```
Meas:Ch1:XParam:Period 1
```

or

```
Meas:Ch1:XParam:Freq?
```

returns either ON or OFF.

### **Group-on mode**

Group-on mode is used to simultaneously turn on a custom group of items. In this case the :Include mnemonic is added to the end of the command. This is then followed by a space and a few items separated by commas. For example, this command turns on the rise time and fall time measurements for Channel 1:

```
Meas:Ch1:XParam:Include Rise,Fall
```

### **Group-off mode**

Group-off mode is used to simultaneously turn off a custom group of items. In this case the :Exclude mnemonic is added to the end of the command. This is then followed by a space and a few items separated by commas. For example, this command turns off the frequency and period measurements for Channel 1:

```
Meas:Ch1:XParam:Exclude Freq,Period
```

### **All-off mode**

All-off mode is used for simultaneously turning off all items. In this case the :ClearAll mnemonic is added to the end of the command. For example, the next command turns off all measurements for Channel 1:

```
Meas:Ch1:XParam:ClearAll
```

Group-on, Group-off and All-off modes do not have a query form.

## Group-query mode

Group-query mode is used find out which items are currently turned on. This mode only has a query form. For example:

```
Meas:Ch1:XParam?
```

The answer may include one or more items separated by a comma, or ClearAll if all items are turned off. For example, the answer Freq,Period means there are two items turned on.

## 3.4 Selector-type commands

The selector-type commands modify a specific instrument setting to one of a few fixed values. For example

```
Trig:Source
```

has these possible arguments:

```
Direct, ExtHF, IntClock
```

and

```
Trig:Mode
```

has these possible arguments:

```
Free, Trig
```

Between 2 and 32 custom arguments are available for these commands. The full set of arguments for each command is specified in the list of commands.

The selector-type commands have a query form. It is possible to use the query form with an argument. For example:

```
Trig:Source? Direct
```

This command sets the Direct input as the trigger source and returns DIRECT.

## 3.5 Integer-type commands

The integer-type commands modify specific integer-value functions. For example, the command

```
Acq:Ch1:RecLen 1024
```

sets the length of Channel 1 signals to 1024 points. The valid range and increment of each value is different and is described in the list of commands.

The integer-type commands have a query form. It is possible to use the query form with an argument. For example

```
Acq:Ch1:RecLen? 24
```

returns 32, since 32 is the minimum valid length of a signal.

### 3.6 Float-type commands

The float-type commands modify specific real-value functions. For example, the command

```
Ch1: Scale 0.1
```

sets the Y-scale for Channel 1 to 100 mV/div. The valid range and increment of each value is different and is described in the list of commands.

Float-type commands have a query form. It is also possible to use the query form with an argument. For example

```
Ch1:Scale? 0.1
```

returns 100 mV/div, when v/div is a dimension of the scale, and the prefix m is milli.

The commands

```
TB:ScaleA? 0.0000001
```

```
TB:ScaleA? 100e-9
```

```
TB:ScaleA? 0.1u
```

```
TB:ScaleA? 100p
```

are equivalent and set the Scale A of the timebase to the value 100 ns/div. All of these commands return 100 ns/div.

### 3.7 Data-type commands

The data-type commands are used to send data to the instrument or to receive data from the instrument, such as the array of points from an acquired signal, the result of a measurement, and so on.

Some data-type commands only have a query form, while others have both a command and a query form. The structure of the data is different for each command and is specified in the list of commands.

## 4 Full list of commands

### 4.1 Header commands

Header: Header

Type: On/Off

Action: Enables/disables headers as part of the query response

### 4.2 GUI commands

#### ***GUI control command***

Header: Gui (SW ver. before 3.20.12)

Header: Gui:Control (SW ver. 3.20.12 or newest)

Type: Selector

Arguments: RemoteLocal, RemoteOnly, Invisible

Action: Set the behavior of the GUI when it is controlled as a COM object.

#### ***GUI side menu command***

Header: Gui:SideMenu:Left:Menu  
Gui:SideMenu:Right:Menu

Type: Selector

Arguments: Off, Ch, Acq, Trig, Displ, Save, Mark, Meas, Math,  
Hist, Eye, Mask, Aux, TDR, Module,  
Util

Action: Remove or Set the specified side menu panel.

#### ***GUI side menu page command***

Header: Gui:SideMenu:Left:Page  
Gui:SideMenu:Right:Page

Type: Integer

Argument: 1 to N, when N is count of pages in the current side menu

Action: Select the page in the specified side menu panel.

Note: This command applies to the side menu with two or more pages.

#### ***GUI side menu signal command***

Header: Gui:SideMenu:Left:Signal  
Gui:SideMenu:Right:Signal

Type: Integer

Argument: 1 to N, when N is count of active signals (max. 4)

Action: Select the signal in the specified side menu panel.

Note: This command applies to the Channels, Save/Recall and Math menus.

## 4.3 System commands

### ***Clear Display***

Header: \*ClrDispl

Type: Execution

Action: Clears the display immediately

### ***Running Control***

Header: \*RunControl

Type: Selector

Arguments: Stop, Single, Run

Action:	Run	- Start a continuous acquisition
	Single	- Start a single acquisition
	Stop	- Immediately stop the acquisition

Response:	Run	- the instrument is in the continuous acquisition state
	Single	- the instrument is in the single acquisition state
	Stop	- the instrument is stopped

### ***Start Autoscaling***

Header: \*Autoscale

Type: Selector

Arguments: Auto, SingleVal, NRZ, RZ

Action: Sets the type of signal and starts autoscaling the instrument

Response: Selected type of signals

### ***Recall Default Setup***

Header: \*DefSetup

Type: Execution

Action: Restores the instrument to its default setup

### ***Set Copy Mode and Copy to the Clipboard***

Header: \*Copy:<Mode>

when <Mode> is one of:

FullScreen	FullWindow
ClientPart	InvClientPart
ScopeScreen	InvScopeScr

Type: Executing-type command

Action: Sets the specified copy mode (All display, software window, client part of the software window, client part of the software window with colors inverted, software screen area or software screen area with color inversion) and copy specified onto the clipboard.

### ***Copy to the Clipboard***

Header: \*Copy

Type: Execution

Action: Puts the image onto the clipboard, depending on the Copy Mode

### ***Get Copy Mode query***

Header: \*Copy?

Argument: None

Forms: Query only

Action: Returns current Copy Mode. See [Set Copy Mode and Copy to the Clipboard](#).

## 4.4 Channels commands

## **Type of the Signals**

Header: AllChs:FitAcqTo

Type: Selector

Arguments: SingleValued, MultiValued

Action: Prepares the instruments for best acquisition of single-valued or multivalued signals

## **Digital Feedback**

Header: AllChs:DigitalFB

Type On/off

Action: Turns digital feedback on or off

## ***Best Flatness***

Header: AllChs:BestFlat

Type On/off

Action: Turns digital correction of low-frequency distortion on or off

### ***Display a Channel***

Header: Ch1:Display Ch2:Display  
Ch3:Display Ch4:Display

Type: On/off

Action: Turns display of the corresponding channel's signal on or off

### ***Acquire a Channel***

Header: Ch1:AcqOnlyEn Ch2:AcqOnlyEn  
Ch3:AcqOnlyEn Ch4:AcqOnlyEn

Type: On/off

Action: On -  
acquisition of the channel is independent of whether it is displayed or  
not  
Off -  
acquisition of the channel occurs only when the channel display is On

**Scale a Channel**

Header: Ch1:Scale Ch2:Scale  
           Ch3:Scale Ch4:Scale

Type: Float

Argument: 0.002 to 0.5, or other when attenuator is used

Action: Sets the specified display scale in V/div

**Scale a TDR Channel**

Header: Ch1:RhoScale Ch2:RhoScale  
           Ch3:RhoScale Ch4:RhoScale (in Rho/div)

Header: Ch1:OhmScale Ch2:OhmScale  
           Ch3:OhmScale Ch4:OhmScale (in Ohm/div)

Type: Float

Argument: 0.01 to 2 (Rho scales)  
           1 to 100 (Ohm scales)

Action: set the specified vertical display scale for TDR non-voltage scales

**Scale a TDT Channel**

Header: Ch1:GainScale Ch2:GainScale  
           Ch3:GainScale Ch4:GainScale (in 1/div)

Header: Ch1:GainDBScale Ch2:GainDBScale  
           Ch3:GainDBScale Ch4:GainDBScale (in dB/div)

Type: Float

Argument: 0.01 to 100 (Gain scales)  
           1 to 20 (Gain dB scales)

Action: set the specified display scale for TDT non-voltage scales

**Offset a Channel**

Header: Ch1:Offset Ch2:Offset  
           Ch3:Offset Ch4:Offset

Type: Float

Argument: -1 to +1, or other when attenuator is used

Action: Sets the specified compensation voltage of the channel in V

**Offset a TDR Channel**

Header: Ch1:RhoOffset Ch2:RhoOffset  
           Ch3:RhoOffset Ch4:RhoOffset (in Rho)

Header: Ch1:OhmOffset  
Ch3:OhmOffset Ch2:OhmOffset  
Ch4:OhmOffset (in Ohm)

Type: Float

Argument: -8 to 8 (Rho units)  
-400 to 400 (Ohm units)

Action:set the specified compensation value for TDR non-voltage scales.

#### **4.4.11 Offset a TDT Channel**

Header: Ch1:GainOffset  
Ch3:GainOffset Ch2:GainOffset  
Ch4:GainOffset (in 1)

Header: Ch1:GainDBOffset  
Ch3:GainDBOffset Ch2:GainDBOffset  
Ch4:GainDBOffset (in dB)

Type: Float

Argument: -400 to 400 (Gain units)  
-80 to 80 (Gain dB units)

Action:set the specified compensation value for TDT non-voltage scales.

#### **Position a Channel**

Header: Ch1:Position  
Ch3:Position Ch2:Position  
Ch4:Position

Type: Float

Argument: -5 to +5

Action: Sets the specified vertical position of the channel on the screen, in divisions.

#### **Bandwidth of Channel**

Header: Ch1:Band  
Ch3:Band Ch2:Band  
Ch4:Band

Type: Selector

Arguments: Full, Narrow

Action: Sets the bandwidth of the channel

#### **Deskew of Channel**

Header: Ch1:Deskew  
Ch3:Deskew Ch2:Deskew  
Ch4:Deskew

Type: Float

Argument: 0 to 100e-9

Action: Sets the deskew of the channel in s

### **Attenuator linear/log**

Header: Ch1:Atten:Unit      Ch2:Atten:Unit  
           Ch3:Atten:Unit      Ch4:Atten:Unit

Type: Selector

Arguments: Off, Ratio, DB

Action: Sets the presence and scale of the attenuator or converter used with the channel

### **Attenuator ratio**

Header: Ch1:Atten:Ratio      Ch2:Atten:Ratio  
           Ch3:Atten:Ratio      Ch4:Atten:Ratio

Type: Float

Argument: 0.0001 to 1000000

Action: Sets the attenuation ratio. This setting is active only when the attenuator unit is *ratio*.

### **Attenuator dB**

Header: Ch1:Atten:DB      Ch2:Atten:DB  
           Ch3:Atten:DB      Ch4:Atten:DB

Type: Float

Argument: -80 to +120

Action: Sets the attenuation in dB. This setting is only active when the attenuator unit is *decibels*.

### **Attenuator unit**

Header: Ch1:Atten:Dimens      Ch2:Atten:Dimens  
           Ch3:Atten:Dimens      Ch4:Atten:Dimens

Type: Selector

Arguments: Volt, Watt, Ampere, Unknown

Action: Sets the units of the converter used with the channel

### **Sampler Delay**

Header: Ch1\_2:SmpDelay      Ch3\_4:SmpDelay

Type: Float

Argument: 0 to 2e-9

Action: Sets the sampler delay in seconds

## 4.5 Timebase commands

### **Timebase Units**

Header: TB:Units

Type: Selector

Arguments: Time, Bit

Action: Sets the units of the timebase to s/div or bit/div

### **Bit rate value**

Header: TB:BitRate

Type: Selector

Arguments: 44M736, 51M84, 125M000, 132M813, 139M264, 155M52, 265M625, 270M000, 393M22, 466M56, 480M000, 531M25, 614M4, 622M08, 786M43, 933M12, 1G0625, 1G2288, 1G24416, 1G25000, 1G485, 1G5000, 1G5729, 1G86625, 2G1231, 2G125, 2G48832, 2G5000, 2G66606, 2G8576, 2G97000, 3G000, 3G072, 3G125, 3G187, 3G25000, 3G32000, 4G25000, 5G000, 6G000, 6G25000, 6G375, 8G5000, 9G95328, 10G3125, 10G51875, 10G6642, 10G709, 11G000, 11G0957, 11G317, 12G2495, 17G000, 19G90656, 25G78125, 27G73193, 39G81312, 42G65691, 43G01841

Action: Sets one of the standard bit rate values from 44.736 Mbit/s to 43.01841 Gbit/s

### **Sampling Mode**

Header: TB:SampleMode

Type: Selector

Arguments: Scope, TDR, Eye, Random, RealTime, Roll

Action: Sets the instrument's sampling mode

### **Timebase mode**

Header: TB:Mode

Type: Selector

Arguments: A, AB, B

Action: Sets main, intensified, or delayed timebase

***Main timebase scale, sec/div***

Header: TB:ScaleA  
 Type: Float  
 Argument: 10e-12 to 10  
 Action: Sets the scale of the main timebase when time units are used

***Delayed timebase scale, sec/div***

Header: TB:ScaleB  
 Type: Float  
 Argument: 10e-12 to 10  
 Action: Sets the scale of delayed timebase when time units are used

***Main timebase scale, bit/div***

Header: TB:BitScaleA  
 Type: Float  
 Argument: 0.0005 to 100000, depending on actual bit rate  
 Action: Sets the scale of the main timebase when bit units are used

***Delayed timebase scale, bit/div***

Header: TB:BitScaleB  
 Type: Float  
 Argument: 0.0005 to 100000, depending on actual bit rate  
 Action: Sets the scale of the delayed timebase when bit units are used

***Main Distances Scale for TDR-TDT***

Header: TB:ScaleMetreA  
 TB:ScaleFootA  
 TB:ScaleInchA  
 Type: Float  
 Argument: depends on actual TDR/TDT dielectric constants or velocity  
 Action: set scale of main timebase when corresponding units are used

***Delayed Distances Scale for TDR-TDT***

Header: TB:ScaleMetreB  
 TB:ScaleFootB  
 TB:ScaleInchB  
 Type: Float

Argument: depends on actual TDR/TDT dielectric constants or velocity

Action: set scale of delayed timebase when corresponding units are used

### ***Timebase delay, s***

Header: TB:Delay

Type: Float

Argument: 0 to max, where max depends on main and delayed timebase

Action: Sets the delay of intensified, delayed timebase when time units are used

### ***Timebase delay, bits***

Header: TB:Delay

Type: Float

Argument: 0 to max, where max depends on main and delayed timebase

Action: Sets the delay of intensified, delayed timebase when bit units are used

### ***Distances Delay for TDR-TDT***

Header: TB:Delay

Type: Float-type command

Argument: 0 to max, where max depends of main and delayed timebase

Action: set delay of intensified, delayed timebase when distances units are used

### ***Real Time Delay***

Header: TB:RTDelay

Type: Float

Argument: 0 to 8

Action: Sets the delay when Random or RealTime sampling modes are used

### ***Real Time Trigger Position***

Header: TB:TrigPos

Type: Float

Argument: 0 to 100

Action: Sets the trigger position when Random or RealTime sampling modes are used in %

## 4.6 Trigger commands

### 4.6.1 Trigger Source

#### ***Trigger Source***

Header: Trig:Source

Type: Selector

Arguments: ExtDirect, ExtPrescaler, Ch1Direct, Ch2Direct,  
IntClock, Auxiliary

### 4.6.2 External Direct Trigger Commands

#### ***External Direct Trigger Level***

Header: Trig:ExtDir:Level

Type: Float

Argument: -1 to +1

Action: Sets the trigger level for external direct input, volts

#### ***External Direct Trigger Slope***

Header: Trig:ExtDir:Slope

Type: Selector

Arguments: Pos, Neg

Action: Sets the Positive or Negative slope of external direct trigger

#### ***External Direct Trigger Hysteresis***

Header: Trig:ExtDir:Hyst

Type: Selector

Arguments: Norm, HighSens

Action: Sets the hysteresis for external direct trigger (Norm) or set off  
(HighSens)

### 4.6.3 External Prescale Trigger Commands

#### ***Automatic external prescaler***

Header: Trig:ExtPresc:AutoNDiv

Type: On/off

Action: Turn the automatic mode of the ext. prescaler on or off

***External prescaler division factor***

Header: Trig:ExtPresc:NDiv  
Type: Selector  
Arguments: 1 to 8  
Forms: Must use the *set with query* or *query* forms only!  
Action: Sets the external prescaler division factor. This query is necessary as the real factor may differ from the arguments.

## 4.6.4 Ch1/Ch2 Direct Trigger Commands

***Ch1/Ch2 Direct Trigger Level***

Header: Trig:Ch1:Level      Trig:Ch2:Level  
Type: Float  
Argument: -1 to +1  
Action: Sets the trigger level in volts for Ch1/Ch2 input

***Ch1/Ch2 Direct Trigger Slope***

Header: Trig:Ch1:Slope Trig:Ch2:Slope  
Type: Selector  
Arguments: Pos, Neg  
Action: Sets the Positive or Negative slope of Ch1/Ch2 input

***Ch1/Ch2 Direct Trigger Hysteresis***

Header: Trig:Ch1:Hyst      Trig:Ch2:Hyst  
Type: Selector  
Arguments: Norm, HighSens  
Action: Sets the hysteresis for Ch1/Ch2 trigger input (Norm) or set off (HighSens)

***Ch1/Ch2 Direct Trigger Coupling***

Header: Trig:Ch1:Coupling      Trig:Ch2:Coupling  
Type: Selector  
Arguments: DC, AC  
Action: Sets the coupling for Ch1/Ch2 trigger input

***Ch1/Ch2 Direct Trigger Rejection***

Header: Trig:Ch1:Reject Trig:Ch2:Reject

Type: Selector

Arguments: Off, LF, HF

Action: Sets the rejection mode for Ch1/Ch2 trigger input

## 4.6.5 Trigger Period for Internal Clock Sources

***Trigger Period for Internal Clock Sources***

Header: Trig:IntRate

Type: Float

Argument: 8e-9 to 0.06

Action: Sets the period for the internal clock trigger source in seconds

## 4.6.6 Trigger Mode

***Trigger Mode***

Header: Trig:Mode

Type: Selector

Arguments: Free, Trig

Action: Sets Freerun or Triggered mode for the trigger

## 4.6.7 Trigger Holdoff Commands

***Holdoff Mode***

Header: Trig:HoldoffBy

Type: Selector

Arguments: Time, Events

Action: Sets the Holdoff mode by time or by events

***Holdoff by Time***

Header: Trig:HoldoffTime

Type: Float

Argument: 1e-6 to 17

Action: Sets the holdoff time in seconds

***Holdoff by Events***

Header: Trig:HoldoffEvents  
Type: Integer  
Argument: 50 to 50000000, depends on actual trigger frequency  
Forms: Must use the *set with query* or *query* forms only  
Action: Sets the holdoff by events. This query is necessary as the real number of events may differ from the arguments.

## 4.6.8 Attenuator Commands

***Attenuator Unit for External Direct Input***

Header: Trig:Atten:Unit  
Type: Selector  
Arguments: Off, Ratio, DB  
Action: Sets the presence and unit of the attenuator or converter used with direct trigger input

***External Direct Input Attenuation (ratio)***

Header: Trig:Atten:Ratio  
Type: Float  
Argument: 0.0001 to 1000000  
Action: Sets the attenuation ratio. This setting is active only when the attenuator unit is ratio.

***External Direct Input Attenuation (dB)***

Header: Trig:Atten:DB  
Type: Float  
Argument: -80 to +120  
Action: Sets the attenuation in dB. This setting is active only when the attenuator unit is decibels.

## 4.6.9 Pattern Sync Trigger Commands

***Pattern Lock Mode***

Header: Trig:PatternLockMode  
Type: Selector  
Arguments: Off, AutoDetect, Manual  
Action: Sets the pattern lock mode

***Pattern Length***

Header: Trig:PattLen  
Type: Integer  
Arguments: 7 to 8388608  
Action: Sets the pattern length

***Start Bit***

Header: Trig:PattDelay  
Type: Integer  
Arguments: 0 to 126  
Action: Sets the pattern delay in bits

***Eye Line Mode***

Header: Trig:PattEyeLine  
Type: On/off  
Action: Turn the eye line mode on or off

***Scan Bits***

Header: Trig:PattScanBits  
Type: Integer-type command  
Arguments: 1 to 127  
Action: Sets the amount of scanning bits in the eye line mode

## 4.7 Acquisition commands

### 4.7.1 Sampling Mode

#### ***Sampling Mode***

Header: Acq:Sampl

Type: Selector

Arguments: Simult, Altern

Action: Simult - sets simultaneous acquisition of all channels  
Altern - sets alternate acquisitions of all channels

### 4.7.2 Channel Acquisition Commands

#### ***Acquisition Mode of Channel***

Header: Acq:Ch1:Mode      Acq:Ch2:Mode  
          Acq:Ch3:Mode      Acq:Ch4:Mode

Type: Selector

Arguments: Sample, AvgStab, AvgMult, EnvMinMax, EnvMax, EnvMin

Action: Sets the acquisition mode of the specified channel

#### ***Channel Averaging***

Header: Acq:Ch1:NAvg      Acq:Ch2:NAvg  
          Acq:Ch3:NAvg      Acq:Ch4:NAvg

Type: Integer

Argument: 1, 2, 4, 8, 16, ... 4096

Action: Sets the averaging coefficient for the specified channel

#### ***Channel Envelope***

Header: Acq:Ch1:NEnv      Acq:Ch2:NEnv  
          Acq:Ch3:NEnv      Acq:Ch4:NEnv

Type: Integer

Argument: 1, 2, 4, 8, 16, ... , 4096, 8192

Action: Sets the number of signals for envelope mode for the specified channel.  
Argument 8192 is used for unlimited number of signals.

### ***Channel Record Length***

Header:      Acq:Ch1:RecLen                    Acq:Ch2:RecLen  
                   Acq:Ch3:RecLen                    Acq:Ch4:RecLen

Type:          Integer

Argument:     32, 64, 128, ... , 432768

Action:        Sets the number of points for specified channel

#### 4.7.3 Termination of the Acquisition

##### ***Termination of Acquisition***

Header:        Acq:RunUntil

Type:           Selector

Arguments:    StopBtn, NAcq

Action:        Sets the condition for terminating acquisition when the Stop button is pressed or after the specified number of waveforms is reached.

#### 4.7.4 Number of Waveforms

##### ***Number of Waveforms***

Header:        Acq:NAcq

Type:           Integer

Argument:     1 to 65535

Action:        Sets the number of signals for the terminating acquisition

#### 4.7.5 Action when Number of Waveforms reached

##### ***Action when Number of Waveforms reached***

Header:        Acq:Action

Type:           On/off-group

Items:        Beep, Save

Action:        If Save is turned on, every signal is stored to disk  
                   If Beep is turned on, the beep signal will sound after the specified number of waveforms is reached

#### 4.7.6 File Name

##### ***File Name***

Header: Acq:FileName

Type: Data

Argument: Text string contains the file path

Action: Defines the full path and base file name for storing the acquired signals onto the Disk. The name of each saved file consists of a base name, followed by an underline (\_) and five-digit auto-incremented numbers. For example:

After the command:

Acq:FileName C:\Temp\Test1\basename

Files basename\_00001.wfm, basename\_00002.wfm,  
basename\_00003.wfm and so on will be written to the C:\Temp  
\Test1 folder.

Note: The specified folder must exist

#### 4.7.7 Stored Files Format

##### ***Stored Files Format***

Header: Acq:FileFormat

Type: Selector

Arguments: Binary, Verbose, YOnly

Action: Sets the format of the file

## 4.8 Display commands

Mnemonic <src> in some Display Commands signifies Source  
( <src> is: Ch1, Ch1B2, Ch2, Ch2B2, F1, F2, F3, F4, M1, M2, M3, M4, S1, S2)

### **Trace mode**

Header: Displ:TraceMode

Type: Selector

Arguments: AllLocked, PerTrace

Action: In PerTrace mode, every waveform may be displayed in its own style  
In AllLocked mode, the display style of all waveforms is set as the style of the active trace

### **Select active trace**

Header: Displ:TraceSel

Type: Selector

Arguments: Ch1, Ch1B2, Ch2, Ch2B2, F1, F2, F3, F4, M1, M2, M3, M4, XY

Action: Selects the active trace for AllLocked trace mode

### **Display Persistence (SW v.3.20 or newer)**

Header: Displ:<src>:InkStyle

Type: Selector

Arguments: Dots, VarPersist, InfinPers, VarGrayScal,  
InfGrayScal,  
VColorGrad, IColorGrad

Action: set display persistence for specified trace in PerTrace mode;  
set display persistence for all traces in AllLocked mode.

Note use argument Dots for disable persistence

### **Display Style (SW v.3.20 or newer)**

Header: Displ:<src>:ChartStyle

Type: Selector

Arguments: Dots, Vectors

Action: set display style for specified trace in PerTrace mode;  
set display style for all traces in AllLocked mode

### **Set Display Style (SW v.3.19 or older)**

Header: Displ:<src>:Style

Type: Selector

Arguments: Dots, Vectors, VarPersist, InfinPers, VarGrayScal, InfGrayScal, VColorGrad, IColorGrad

Action: Sets the display style for specified trace in PerTrace mode  
Sets the display style for all traces in AllLocked mode

#### ***Persistence Time , seconds (for VarPersist Style)***

Header: Displ:<src>:PersistTime

Type: Float

Argument: 0.1 to 20

Action: Sets the persistence time for specified trace in PerTrace mode  
Sets the persistence time for all traces in AllLocked mode

#### ***Refresh Time, seconds (for VarGrayScal or VColorGrade Styles)***

Header: Displ:<src>:RefreshTime

Type: Float

Argument: 1 to 200

Action: Sets the refresh time for specified trace in PerTrace mode  
Sets the refresh time for all traces in AllLocked mode

#### ***Reset Display Style***

Header: Displ:ResetAll

Type: Execution

Action: Resets Display Styles to initial state (variable persistence 2 c)

#### ***Display Format***

Header: Displ:Format

Type: Selector

Arguments: Auto, YT, 2YT, 4YT, XY, CombYTXY, Comb2YTXY

Action: Selects the number and kinds of screens

#### ***Define Trace Screen (for 4YT Format)***

Header: Displ:Screen4:<trace>,  
when <trace> is <src> or Hist

Type: Selector

Arguments: 1, 2, 3, 4

Action: Moves the specified trace onto the specified screen in 4YT format

***Define Trace Screen (for 2YT, Comb2YTXY Formats)***

Header: Displ:Screen2:<trace>,  
when <trace> is <src> or Hist

Type: Selector

Arguments: 1, 2

Action: Moves the specified trace onto the specified screen in 2YT or Comb2YTXY formats

***Source of X Axis for XY Screen***

Header: Displ:XAxis

Type: Selector

Arguments: <src>, exclude XY, DB

Action: Sets the specified signal as X axis for XY screen

***Source of Y Axis for XY Screen***

Header: Displ:YAxis

Type: Selector

Arguments: <src>, exclude XY, DB

Action: Sets the specified signal as Y axis for XY screen

***Graticule Type***

Header: Displ:Gratic

Type: Selector

Arguments: Full, Frame, Axis, Off

Action: Defines the type of graticule for YT and XY screens

## 4.9 Save/Recall commands

### 4.9.1 Work with Memo Zones (M1, M2, M3, M4)

#### ***Memory Display***

Header: Save:<mz>:Visible

Type: On/off-group

Items: M1, M2, M3, M4

Action: Controls the display of memory zones

#### ***Source for storing into Memory***

Header: Save:Memo:Source

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Defines the signal as source for storing into memory zone

#### ***Save into Memory***

Header: Save:<mz>:Save

Type: Execution

Action: Stores the selected source into selected memory

### 4.9.2 Memory Scaling

#### ***Complex Format***

Header: Save:<mz>:ComplexScale

Type: Selector

Arguments: Magnitude, Phase, Real, Imaginary

Action: Defines the display mode when memory contains spectrum

#### ***Vertical Scale Type***

Header: Save:<mz>:VScaleType

Type: Selector

Arguments: Linear, Logarithm

Action: Defines the vertical scale type for the magnitude of the spectrum

***Vertical linear Scale***

Header: Save:<mz>:VertScale  
Type: Float  
Arguments: 1e-6 to 1e6  
Action: Defines the vertical scale in volts/div for Linear vertical scale type

***Vertical linear Position***

Header: Save:<mz>:VoltPosit  
Type: Float  
Arguments: -10 to +10  
Action: Defines the vertical position in div for Linear vertical scale type

***Vertical logarithmic Scale***

Header: Save:<mz>:VertDBScale  
Type: Float  
Arguments: 1 to 120  
Action: Defines the vertical scale in dB/div for Logarithm vertical scale type

***Vertical logarithmic Position***

Header: Save:<mz>:Posit0DB  
Type: Float  
Arguments: -10 to +10  
Action: Defines the vertical position in div for Logarithm vertical scale type

***Vertical Phase Scale***

Header: Save:<mz>:PhaseScale  
Type: Float  
Arguments: 0.125 to 8  
Action: Defines the vertical scale in rad/div for Phase display mode

***Vertical Phase Position***

Header: Save:<mz>:PhasePosit

Type: Float

Arguments: -10 to +10

Action: Defines the vertical position in div for Phase display mode

***Horizontal Scale***

Header: Save:<mz>:HorScale

Type: Float

Arguments: Depends on actual timebase

Action: Defines the horizontal scale in actual X-axis unit/div

***Horizontal Position***

Header: Save:<mz>:HorPosition

Type: Float

Arguments: Depends on actual timebase

Action: Defines the horizontal position in actual X-axis unit

#### 4.9.3 Work with Disk

***File Type***

Header: Save:Disk:FileType

Type: Selector

Arguments: Wfm, DB

Action: Defines the file type for saving

***Source for saving to file***

Header: Save:Disk:Source

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Defines the signal as source for saving to file

**File Name**

Header: Save:Disk:FileName  
Type: Data  
Argument: Text string  
Forms: Command, query, command with query  
Action: Defines the file name for saving the specified signal to disk  
Note: Specified folder must exist

**File Name Mode**

Header: Save:Disk:NameMode  
Type: Selector  
Arguments: Manual, Auto  
Action: Sets the file name mode. In Auto mode the file name consists of a base name followed by an underscore (\_) and a five-digit number. Each time you save a waveform, the number in the file name is automatically incremented. For example: basename\_00001.wfm, basename\_00002.wfm, basename\_00003.wfm, and so on.

**Format of stored files**

Header: Save:Disk:FileFormat  
Type: Selector  
Arguments: Binary, Verbose, YOnly  
Action: Sets the file format

**Save to Disk**

Header: Save:Disk:ExecSave  
Type: Executing  
Action: Saves the selected source to previously specified file

**Load from Disk**

Header: Save:<mz>:LoadFromDsk  
Type: Executing  
Action: Loads the previously specified disk file into the specified Memory Zone

#### 4.9.4 Work with Setups

##### ***Recall Factory Setup***

Header: Save:Setup:RecFact  
Type: Execution  
Action: Returns the instrument to manufacturer's default setting

##### ***Recall Default Setup***

Header: Save:Setup:RecDefault  
Type: Executing  
Action: Returns the instrument to its default setting

##### ***Recall Power-Off Setup***

Header: Save:Setup:RecLast  
Type: Execution  
Action: Returns the instrument to the last setting before the power supply was last switched off

##### ***Save Setup as Default***

Header: Save:Setup:SvAsDefault  
Type: Execution  
Action: Stores the present front-panel setup as the default setup

##### ***Name of Custom Setup File***

Header: Save:Setup:FileName  
Type: Data  
Argument: Text string containing file path  
Action: Defines the file name for storing Custom Setup  
Note: The specified folder must exist

##### ***Save Custom Setup***

Header: Save:Setup:Save  
Type: Execution  
Action: Stores the present front-panel setup as previously specified custom setup

***Recall Custom Setup***

Header: Save:Setup:Recall

Type: Execution

Action: Recalls the setup previously saved to file. The name of the setup must first be defined by the command Save:Setup:FileName .

## 4.10 Markers commands

### ***Marker Type***

Header:      `Mark:Type`  
 Type:        Selector  
 Arguments:    Off, MX, MY, XY  
 Action:       Sets the marker type

### ***Marker Sources***

Header:      `Mark:M1:Source`                  `Mark:M2:Source`  
 Type:        Selector  
 Arguments:    Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4  
 Action:       Attaches the specified marker to the specified signal

### ***X position of Marker***

Header:      `Mark:M1:XPos`                  `Mark:M2:XPos`  
 Type:        Float  
 Argument:    Real value of X-axis  
 Action:       Sets the X position of the specified marker

### ***Y position of Marker***

Header:      `Mark:M1:YPos`, `Mark:M2:YPos`  
 Type:        Float  
 Argument:    Real value of Y-axis  
 Action:       Sets the Y position of the specified marker

### ***Y value of XY-Marker query***

Header:      `Mark:M1:YVal?`                  `Mark:M2:YVal?`  
 Type:        Float  
 Argument:    none  
 Forms:       query only  
 Action:       return value of the signal in the XY-marker point.  
 Note:        return zero when marker type is Off, MX;  
               return Y position of marker when marker type is MY

### **Motion of Markers**

Header: Mark:Motion

Type: Selector

Arguments: Independ, Paired

Action: When Paired motion is selected, you can move both markers with the M1 POSITION variable simultaneously, while the difference between markers can be moved with the M2 POSITION variable

## 4.11 Measure commands

### 4.11.1 Common Measures commands

The mnemonic <src> in some Measure Commands signifies the Source (<src> is: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4)

#### **Measurement Type**

Header: Meas:Display

Type: Selector

Arguments: Off, Param, Statistic

Action: Sets the measurement type

#### **Measurement Source**

Header: Meas:DisplSrc

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Sets the source for the measurement

#### **Measurement Mode**

Header: Meas:Mode

Type: Selector

Arguments: Permanent, Single

Action: Sets the measurement mode

#### **Execute Single Measurement**

Header: Meas:SingleMeas

Type: Execution

Action: Executes a single measurement in Single mode

#### 4.11.2 Statistic Commands

##### **Statistic Measurement Mode**

Header: Meas:StatMode

Type: Selector

Arguments: Permanent, Window, Weight

Action: Sets the Statistic Measurement mode

##### **Windows Value**

Header: Meas:Window

Type: Integer

Argument: 8 to 8192

Action: Sets the number of recently acquired waveforms for Window mode of Statistic Measurement

##### **Weight Value**

Header: Meas:Weight

Type: Integer

Argument: 8 to 8192

Action: Sets the weight variable for Weight mode of Statistic Measurement

##### **Clear Statistics for all Measures**

Header: Meas:ClrStatistics

Type: Executing-type command

Action: Clear statistics of all measurements for all signals

#### 4.11.3 Define parameter Commands

##### **Viewing of Define Parameters**

Header: Meas:View

Type: On/off-type command

Action: Sets the visibility of *define parameters* markers for selected sources

##### **Top/Base Definition Method**

Header: Meas:<src>:Method

Type: Selector

Arguments: Hist, MinMax, Marker

Action: Sets the Top and Base vertical reference thresholds for amplitude measurements of specified signals

### ***Top Value for Marker Method***

Header: Meas:<src>:Top

Type: Integer

Argument: 257 to 1023

Action: Sets the Top vertical reference threshold for specified signals. Argument 0 corresponds to the bottom of the screen, and argument 1023 corresponds to the top of the screen independently of the real screen's height.

### ***Base Value (for Marker Method)***

Header: Meas:<src>:Base

Type: Integer

Argument: 1 to 767

Action: Sets the Base vertical reference threshold for specified signals. Argument 0 corresponds to the bottom of the screen, and argument 1023 corresponds to the top of the screen independent of the real screen's height.

### ***Threshold Definition Method***

Header: Meas:<src>:Thresh

Type: Selector

Arguments: 10-90, 20-80, Custom

Action: Sets the lower, middle, and upper thresholds for measurements of the specified signals. May be set to the fixed values 10%-50%-90%; 20%-50%-80%; or custom values.

### ***Threshold Units***

Header: Meas:<src>:Unit

Type: Selector

Arguments: Percent, Volt, Division

Action: Sets the units of thresholds for the specified signals. Used for custom threshold definition method only.

***Position of Upper, Middle or Lower Threshold***

Headers: Meas:<src>:UpThresh  
Meas:<src>:MidThresh  
Meas:<src>:LowThresh

Type: Float

Arguments: Absolute voltage value (for Volt threshold units only)  
-4 to +4 (for Division threshold units only)

Action: Sets the threshold position for the specified signals

***Percentage of Upper, Middle or Lower Threshold***

Headers: Meas:<src>:UpThPerc  
Meas:<src>:MidThPerc  
Meas:<src>:LowThPerc

Type: Integer

Arguments: -80 to +200

Action: Sets the threshold percentage for the specified signals. Used for Percent threshold units only. Argument 0 (%) corresponds to the Base of the signals, and argument 100 (%) corresponds to the Top of the signals.

***Margins Definition Mode***

Header: Meas:<src>:MargMode

Type: Selector

Arguments: Slope, Marker

Action: Sets the margins definition mode

***Slope of Left or Right Margins***

Headers: Meas:<src>:LeftSlope  
Meas:<src>:RightSlope

Type: Integer

Arguments: 0 to 127

Action: Sets the margin for the specified signals on the specified slope. Used for slope margins definition mode only. Argument 0 = the first rise, value 1 = first fall, 2 = second rise, 3 = second fall, and so on.

### ***Thresholds of Left and Right Margin Slopes***

Headers: Meas:<src>:LeftTresh  
Meas:<src>:RightTresh

Type: Selector

Arguments: Upper, Middle, Lower

Action: Sets the thresholds for definitions of the left or right slope. Used for slope margins definition mode only.

### ***Position of Left or Right Margin***

Headers: Meas:<src>:LeftMarker  
Meas:<src>:RightMarker

Type: Float

Arguments: Absolute time value

Action: Sets the position of margin for the specified signals. Used for marker margins definition mode only.

#### 4.11.4 List of Measurements

##### ***List of X Measurements***

Header: Meas:<src>:XParam

Type: On/off-group

Items: Period, Freq, PosWidth, NegWidth, Rise, Fall, PosDuty, NegDuty, PosCross, NegCross, BurstWidth, Cycles, TimeOfMax, TimeOfMin, PosJitterPp, PosJitterRMS, NegJitterPp, NegJitterRMS

Action: Defines the set of X-axis measurements for the specified signals

##### ***List of Y Measurements***

Header: Meas:<src>:YParam

Type: On/off-group

Items: Max, Min, PP, Top, Base, Ampl, Middle, Mean, dcRMS, acRMS, Area, CycMean, CycDcRMS, CycAcRMS, CycArea, PosOver, NegOver

Action: Defines the set of Y-axis measurements for the specified signals

#### 4.11.5 Inter-signal Measurements

##### ***Second Source for Inter-Signal Measurements***

Header: Meas:Source2

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Sets the second source for the inter-signal measurements

##### ***List of X Inter-Signal Measurements***

Header: Meas:<src>:XDualPar

Type: On/off-group

Items: DellR1R, DellR1F, DellF1R, DellF1F, DellRnR,  
DellRnF, DellFnR, DellFnF, PhaseDeg, PhaseRad,  
PhasePerc, Gain, DBGain

#### 4.11.6 Measurements of Spectrum Signals

Mnemonic <fft\_src> in some Measurement Commands signifies Source (<fft\_src> is F1, F2, F3, F4, M1, M2, M3, M4)

##### 4.11.6.1 Spectrum Parameter Commands

##### ***Limits Definition Method for Spectrum***

Header: Meas:<src>:FFTMethod

Type: Selector

Arguments: Harmonic, Peak

Action: Sets the method of the limits definition for the specified signal. Used for spectrum signals only.

### **Left and Right Spectrums Margin**

Headers: Meas:<src>:FFTLeft  
Meas:<src>:FFTRight

Type: Float

Arguments: Absolute frequency value

Action: Sets the position of margin for the specified spectrum signals. Used for searching for peak 1 of the spectrum for the Harmonic method.

### **Peak Level of Spectrum**

Header: Meas:<src>:PeakLevel

Type: Float

Arguments: -100 to +80 (dBV)

Action: Sets the level for the specified spectrum signals. Used for searching a peak of the spectrum for the Peak method.

### **Left and Right Spectrum Peaks**

Headers: Meas:<src>:PeakLeft  
Meas:<src>:PeakRight

Type: Integer

Arguments: 1 to 41

Action: Sets the first and second peaks for the specified spectrum signals

#### 4.11.6.2 List of Spectrum Frequency Measurements

### **List of Spectrum Frequency Measurements**

Header: Meas:<src>:XFFTPar

Type: On/off-group

Items: Freq, DFreq

Action: Defines the set of the frequency measurements for the specified signals

#### 4.11.6.3 List of Spectrum Magnitude Measurements

### **List of Spectrum Magnitude Measurements**

Header: Meas:<src>:YFFTPar

Type: On/off-group

Items: Magn, DMagn, TDH

Action: Defines the set of the magnitude measurements for the specified signals

#### 4.11.7 Delete all Measurements for all Sources

##### **Delete all Measurements for all Sources**

Header: Meas:ClearAll

Type: Execution

Action: Clears the list of all measurements for all signals

#### 4.11.8 Getting Measurement Results

##### **Get List of Measured Parameters**

Header: Meas:Res>List?

Type: Data

Argument: None

Forms: Query only

Action: Returns text with the list of the active measurements for all signals with ordinal index

##### **Get Current Value of Parameter**

Header: Meas:Res:<N>?

Parameter <N>: Index of the parameter in the list

Type: Data

Argument: None

Forms: Query only

Action: Returns the last result of the specified measured parameter

##### **Get Statistic Value of Parameter**

Header: Meas:Res:<N>:<Val>?

Parameter <N>: Index of the parameter in the list

Parameter <Val>: Wfm, Min, Max, Mean, StdDev

Type: Data

Arguments: None

Forms: Command with query only

Action: Returns the specified statistic parameter of the measured parameter

## 4.12 Limit Tests commands

### 4.12.1 Limit Test On/Off

#### ***Limit Test On/Off***

Header: Limit:TestOn

Type: On/off

Action: Enables/disables the Limit Test. Must be set On after full definition of all other Limit Test parameters.

### 4.12.2 Limit Test Termination Commands

#### ***Limit Test Termination Condition***

Header: Limit:RunUntil

Type: Selector

Arguments: StopBtn, Failur, Wfm

Action: Sets the condition of Limit Test Termination

#### ***Number of Failures***

Header: Limit:Failures

Type: Integer

Argument: 1 to 10000

Action: Sets number of failures for the Failur Condition of the Limit

#### ***Number of Waveforms***

Header: Limit:NWfms

Type: Integer

Argument: 1 to 1000000

Action: Sets the number of waveforms for the Wfm Condition of the Limit

### 4.12.3 Limit Test Action Commands

#### ***Action***

Header: Limit:Action

Type: On/off-group

Items: Beep, Save, Stop

Action: Save – every signal with a limit condition is stored to the disk

Beep – the beep signal will sound for every limit condition

Stop – acquisition immediately stops after the first limit condition

**Action If**

Header: Limit:If

Type: Selector

Arguments: AnyFail, AllPass, AllFail, AnyPass

Action: Define the limit condition:  
 AnyFail – one or more active measures fails  
 AllPass – all active measures are good  
 AllFail – all active measures fail  
 AnyPass – one or more active measurements is good

**Format of Stored Files**

Header: Limit:FileFormat

Type: Selector

Arguments: Binary, Verbose, YOnly

Action: Sets the file format

**File Name**

Header: Limit:FileName

Type: Data

Argument: Text string

Forms: Command, query, command with query

Action: Defines the file name for saving the specified signals to disk

## 4.12.4 Parameter Definition Commands

**Parameter Activity**

Headers: Limit1:Activv Limit2:Activv  
 Limit3:Activv Limit4:Activv

Type: On/off

Action: Enables/disables the Limit Test for relevant parameter

**Parameter Limit Mode**

Headers: Limit1:Mode Limit2:Mode  
 Limit3:Mode Limit4:Mode

Type: Selector

Arguments: Center, Limit

Action: Sets the mode of limits for the relevant parameter

### **Upper and Lower Limits of Parameters**

Headers: Limit1:UpLimit      Limit1:LowLimit  
           Limit2:UpLimit      Limit2:LowLimit  
           Limit3:UpLimit      Limit3:LowLimit  
           Limit4:UpLimit      Limit4:LowLimit

Type:      Float

Arguments: Absolute value of limit

Action:     Sets the limit's value. Used only for Limit mode of the parameter's limit.

### **Parameter Center Mode**

Headers: Limit1:CenterMode      Limit2:CenterMode  
           Limit3:CenterMode      Limit4:CenterMode

Type:      Selector

Arguments: CurrMean, UserDef

Action:     Sets the mode of the center definition for the relevant parameter. Used only for the Center mode of the parameter limit.

### **Center Value**

Headers: Limit1:CenterVal      Limit2:CenterVal  
           Limit3:CenterVal      Limit4:CenterVal

Type:      Float

Arguments: Absolute value of center

Action:     Sets the absolute center value. Used for UserDef mode of the center of the parameter.

### **Parameter Delta Mode**

Headers: Limit1:Delta      Limit2:Delta  
           Limit3:Delta      Limit4:Delta

Type:      Selector

Arguments: StdDev, UserDef, UserPerc

Action:     Sets the mode of delta definition for relevant parameter. Used for Center mode of parameter limit only.

### **Parameter Delta Value for Standard Deviation mode**

Headers: Limit1:StdDev      Limit2:StdDev  
           Limit3:StdDev      Limit4:StdDev

Type: Float

Arguments: 0.1 to 100 standard deviations of the parameter

Action: Sets the delta value. Used for StdDev mode of parameter delta only.

### **Parameter Delta Value for User Defined Mode**

Headers: Limit1:UserDef      Limit2:UserDef  
           Limit3:UserDef      Limit4:UserDef

Type: Float

Arguments: Absolute value of delta

Action: Sets the delta value. Used for UserDef mode of delta of the parameter only.

### **Parameter Delta Percentage for User Defined mode**

Headers: Limit1:UserPerc      Limit2:UserPerc  
           Limit3:UserPerc      Limit4:UserPerc

Type: Float

Arguments: 0.01% to 90% standard deviations of the parameter

Action: Sets the delta value. Used for UserPerc mode of delta of the parameter only.

### **Failure When**

Headers: Limit1:FailWhen      Limit2:FailWhen  
           Limit3:FailWhen      Limit4:FailWhen

Type: Selector

Arguments: Outside, Inside, Always

Action: Sets the mode of the quality control for the according parameter

### **If Measurement Undefined**

Headers: Limit1:NotFound      Limit2:NotFound  
           Limit3:NotFound      Limit4:NotFound

Type: Selector

Arguments: Ignore, Fail, Pass

Action: Sets the limit status when measurement is undefined

## 4.13 Mathematics commands

### 4.13.1 Enable Mathematical Function

#### ***Enable Mathematical Function***

Headers: F1:On F2:On  
F3:On F4:On

Type: On/off

Action: Enables/disables the calculation and display of the relevant functions

### 4.13.2 DisplayMathematical Function

#### ***DisplayMathematical Function***

Headers: F1:Display F2:Display  
F3:Display F4:Display

Type: On/off

Action: Enables/disables the visibility of the relevant functions

### 4.13.3 Function Category

#### ***Function Category***

Headers: F1:Category F2:Category  
F3:Category F4:Category

Type: Selector

Arguments: Arithm, Algebra, Trigonom, FFT, BitOp, Misc,  
Formula

Action: Sets the category of the specified function

### 4.13.4 Function Operators

#### ***Arithmetic Function Operator***

Headers: F1:ArithmOp F2:ArithmOp  
F3:ArithmOp F4:ArithmOp

Type: Selector

Arguments: Add, Subtract, Multiply, Divide, Ceil,  
Floor, Fix, Round, Absolute, Invert,  
Common, ReScale

Action: Sets the operator of the specified function. Used for Arithm category only.

***Algebraic Function Operator***

Headers: F1:AlgebraOp                    F2:AlgebraOp  
           F3:AlgebraOp                    F4:AlgebraOp

Type: Selector

Arguments: ExpE, LogE, Exp10, Log10, ExpA, LogA,  
           Differentiate, Integrate, Square, SqRoot,  
           Cube, PowerA, Inverse, SqRtOfSum

Action: Sets the operator of the specified function. Used for Algebra category only.

***Trigonometric Function Operator***

Headers: F1:TrigonOp                    F2:TrigonOp  
           F3:TrigonOp                    F4:TrigonOp

Type: Selector

Arguments: Sine, ASine, Cosine, ACosine, Tangent,  
           ATangent, Cotangent, ACotangent, HSine,  
           HCosine, HTangent, HCotangent

Action: Sets the operator of the specified function. Used for Trigonometric category only.

***FFT Function Operator***

Headers: F1:FFTOp                        F2:FFTOp  
           F3:FFTOp                        F4:FFTOp

Type: Selector

Arguments: FFT, IFFT, FTMagn, FFTPhase, FFTReal, FFTIm

Action: Sets the operator of the specified function. Used for FFT category only.

***Bits Function Operator***

Headers: F1:BitOp                        F2:BitOp  
           F3:BitOp                        F4:BitOp

Type: Selector

Arguments: And, NAnd, Or, NOR, XOr, NXOr, Not

Action: Sets the operator of the specified function. Used for BitOp category only.

### **Miscellaneous Function Operator**

Headers: F1:MiscOp F2:MiscOp  
           F3:MiscOp F4:MiscOp

Type: Selector

Arguments: LinInterp, SinXInterp, Trend, Smooth

Action: Sets the operator of the specified function. Used for Misc category only.

#### 4.13.5 Function Operands

##### **Operand 1**

Headers: F1:Source1 F2:Source1  
           F3:Source1 F4:Source1

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4B2, F1, F2, F3, F4, M1,  
           M2, M3, M4

Action: Sets the first operand of the specified function

##### **Operand 2**

Headers: F1:Source2 F2:Source2  
           F3:Source2 F4:Source2

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1,  
           M2, M3, M4, Constant

Action: Sets the second operand of the specified function. Used for dual- or quad-operand function.

##### **Operands 3/4**

Headers: F1:Source3 F1:Source4  
           F2:Source3 F2:Source4  
           F3:Source3 F3:Source4  
           F4:Source3 F4:Source4

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4,  
           M1, M2, M3, M4, "Don'tCare"

Action: Sets the third and fourth operands for the specified function. Used for bits function.

***Constant Value***

Headers: F1:Const                  F2:Const  
              F3:Const                  F4:Const

Type:        Float

Arguments: Absolute value of constant

Action:      Sets the constant for the specified function. Used when Const is the second operand.

## 4.13.6 Additional Parameters for Arithmetic Functions

***Rounding Step***

Headers: F1:RoundTo                  F2:RoundTo  
              F3:RoundTo                  F4:RoundTo

Type:        Float

Arguments: Value of rounding step

Action:      Sets the step for rounding function. Used for Ceil, Floor, Fix, Round arithmetic functions.

***Rescale Parameters***

Headers: F1:ResMult                  F1:ResOffset  
              F2:ResMult                  F2:ResOffset  
              F3:ResMult                  F3:ResOffset  
              F4:ResMult                  F4:ResOffset

Type:        Float

Arguments: Value of Mult and Offset parameters

Action:      Sets the Mult and Offset parameters. Used for ReScale arithmetic function.

#### 4.13.7 Additional Parameters for Algebraic Functions

##### ***Logarithmic Base***

Headers: F1:LogBase F2:LogBase  
           F3:LogBase F4:LogBase

Type: Float

Arguments: 1.01 to 100

Action: Sets the logarithmic base for LogA algebraic function

##### ***Number Exponent***

Headers: F1:PowerExp F2:PowerExp  
           F3:PowerExp F4:PowerExp

Type: Float

Arguments: -100 to +100

Action: Sets the Number Exponent for ExpA algebraic function

#### 4.13.8 Additional Parameters for Trigonometric Functions

##### ***Volt-to-Radian Coefficient***

Headers: F1:YScaleRad F2:YScaleRad  
           F3:YScaleRad F4:YScaleRad

Type: Float

Arguments: -100 to +100

Action: Sets the volt-to-radian coefficient for att trigonometric functions

#### 4.13.9 Additional Parameters for FFT Functions

##### ***Window***

Headers: F1:Window F2:Window  
           F3:Window F4:Window

Type: Selector

Arguments: Rectang, Hamming, Hanning, FlatTop, BlackHarr,  
                  KaiserBess

Action: Sets the window for the specified function

## ***Suppression***

Headers: F1:Suppress F2:Suppress  
F3:Suppress F4:Suppress

Type: Group-on/off

Arguments: DC, PHASE

Action: DC - on/off the suppression of the spectrum DC component;  
PHASE - on/off the suppression of the spectrum phase noise.

F1:SupprLevel: Float (-120 to -10, dB)

## ***Phase Suppression Level***

Headers: F1 : SupprLevel  
          F3 : SupprLevel

Type: Float

Arguments: -120 dB to -10 dB

Action: Sets the phase suppression level with respect to a maximum magnitude

#### 4.13.10 Additional Parameters for Bit Functions

## **Source Thresholds**

Headers:	F1:Thresh1 F3:Thresh1	F2:Thresh1 F4:Thresh1
	F1:Thresh2 F3:Thresh2	F2:Thresh2 F4:Thresh2
	F1:Thresh3 F3:Thresh3	F2:Thresh3 F4:Thresh3
	F1:Thresh4 F3:Thresh4	F2:Thresh4 F4:Thresh4
Type:	Float	
Arguments:	Value of thresholds	
Action:	Sets the threshold levels for each source of the bit functions	

## **Source Inversion**

Headers: F1:SorceInvert F2:SorceInvert  
F3:SorceInvert F4:SorceInvert

Type: Group-on/off

Arguments: SRC1, SRC2, SRC3, SRC4

Action: Enables/disables the inversion of each source

#### 4.13.11 Additional Parameters for Miscellaneous Functions

##### ***Smoothing Parameter***

Headers: F1:SmoothLen                    F2:SmoothLen  
             F3:SmoothLen                    F4:SmoothLen

Type: Integer

Argument: 0 to 24

Action: Sets the length of the smoothing interval in points for the specified function. Used for Smooth operator only. Length is defined as  $3 + <\text{Argument}> * 2$ .

##### ***Signal Length***

Headers: F1:SignalLen                    F2:SignalLen  
             F3:SignalLen                    F4:SignalLen

Type: Integer-type command

Argument: 2048, 4096 or 8192

Action: Sets the length of the interpolation function signal. Used for LinInterp and SinXInterp functions.

##### ***Trend Measurement***

Headers: F1:TrendMeas                    F2:TrendMeas  
             F3:TrendMeas                    F4:TrendMeas

Type: Selector

Arguments: Period, Freq, PosWidth, NegWidth, RiseTime,  
              FallTime, PosDuty, NegDuty

Action: Sets the kind of trends for the specified function. Used for Trend operator only.

#### 4.13.12 Function Scaling

##### **Complex Format**

Header: F1:ComplexScale      F2:ComplexScale  
          F3:ComplexScale      F4:ComplexScale

Type: Selector

Arguments: Magnitude, Phase, Real, Imaginary

Action: Defines the spectrum display mode for FFT function

##### **Vertical Scale Type**

Header: F1:VScaleType      F2:VScaleType  
          F3:VScaleType      F4:VScaleType

Type: Selector

Arguments: Linear, Logarithm

Action: Defines the vertical scale type for Magnitude of the FFT function

##### **Vertical linear Scale**

Header: F1:VertScale      F2:VertScale  
          F3:VertScale      F4:VertScale

Type: Float

Arguments: 1e-6 to 1e6

Action: Defines the vertical scale in volts/div for Linear vertical scale type

##### **Vertical linear Position**

Header: F1:VoltPosit      F2:VoltPosit  
          F3:VoltPosit      F4:VoltPosit

Type: Float

Arguments: -10 to +10

Action: Defines the vertical position in div for Linear vertical scale type

##### **Vertical logarithmic Scale**

Header: F1:VertDBScale      F2:VertDBScale  
          F3:VertDBScale      F4:VertDBScale

Type: Float

Arguments: 1 to 120

Action: Defines the vertical scale in dB/div for Logarithm vertical scale type

### ***Vertical logarithmic Position***

Header: F1:Posit0DB F2:Posit0DB  
           F3:Posit0DB F4:Posit0DB

Type: Float

Arguments: -10 to +10

Action: Defines the vertical position in div for Logarithm vertical scale type

### ***Vertical Phase Scale***

Header: F1:PhaseScale F2:PhaseScale  
           F3:PhaseScale F4:PhaseScale

Type: Float

Arguments: 0.125 to 8

Action: Defines the vertical scale in rad/div for Phase display mode

### ***Vertical Phase Position***

Header: F1:PhasePosit F2:PhasePosit  
           F3:PhasePosit F4:PhasePosit

Type: Float

Arguments: -10 to +10

Action: Defines the vertical position in div for Phase display mode

### ***Horizontal Scale***

Header: F1:HorScale F2:HorScale  
           F3:HorScale F4:HorScale

Type: Float

Arguments: Depends on actual timebase

Action: Defines the horizontal scale in actual X-axis unit/div

### ***Horizontal Position***

Header: F1:HorPosition F2:HorPosition  
           F3:HorPosition F4:HorPosition

Type: Float

Arguments: Depends on actual timebase

Action: Defines the horizontal position in actual X-axis unit

## 4.14 Histogram commands

### 4.14.1 General Histogram Commands

#### ***Histogram Axis***

Header: Hist:Axis

Type: Selector

Arguments: Off, Vert, Horiz

Action: Sets the axis of the histogram

#### ***Histogram Source***

Header: Hist:Source

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Selects the specified signal as source of the histogram

#### ***Histogram Visibility***

Header: Hist:Visible

Type: On/off

Action: Sets the visibility of the histogram. The acquisition of the histogram proceeds independently of this command.

### 4.14.2 Histogram Completion Commands

#### ***Histogram Finish Condition***

Header: Hist:RunUntil

Type: Selector

Arguments: StopSingle, Wfms, Samples

Action: Sets the finish condition for acquiring the histogram

#### ***Number of Waveforms for Histogram***

Header: Hist:NWfm

Type: Integer-type command

Argument: 1 to 1000000

Action: Sets the number of signals for the termination of histogram acquisition

### **Number of Samples for Histogram**

Header: Hist:NSample  
 Type: Integer-type command  
 Argument: 1 to 10000000  
 Action: Sets the number of samples for the termination of histogram acquisition

#### 4.14.3 Histogram Window Commands

##### **Limit Mode for Histogram Window**

Header: Hist:Limits  
 Type: Selector  
 Arguments: Paired, Independ  
 Action: Sets the mode of the limits of the histogram window

##### **Limit Unit for Histograms Window**

Header: Hist:Units  
 Type: Selector  
 Arguments: Absolute, Percent  
 Action: Sets the units of the limits of the histogram window

##### **Left and Right Window Limits for Vertical or Horizontal Histogram**

Headers: Hist:WVert:Left      Hist:WVert:Right  
           Hist:WHor:Left      Hist:WHor:Right  
 Type: Float  
 Argument: Real value of the X-axis (for Absolute units)  
               0% to 100% of the X-axis (for Percent units)  
 Action: Sets the X positions of the histogram window

##### **Top and Bottom Window Limits for Vertical or Horizontal Histogram**

Headers: Hist:WVert:Top      Hist:WVert:Bottom  
           Hist:WHor:Top      Hist:WHor:Bottom  
 Type: Float  
 Argument: Real value of the Y-axis (for Absolute units)  
               0% to 100% of the Y-axis (for Percent units)  
 Action: Sets the Y positions of the histogram window

**Window Visibility**

Header: Hist:Display  
Type: On/off  
Action: Sets the visibility of the window

**Set Default Window**

Header: Hist:SetDefWind  
Type: Executing-type command  
Action: Sets the default window depending on the axis

## 4.14.4 Histogram Calculation Commands

**Calculation Mode**

Header: Hist:Mode  
Type: Selector  
Arguments: Normal, Exponent  
Action: Sets the mode of histogram calculation

**Weight for Exponential Calculation**

Header: Hist:Weight  
Type: Integer-type command  
Argument: 8, 16, 32, ..., 8192  
Action: Sets the weight coefficient for the Exponent calculation mode

## 4.14.5 Histogram Scale Commands

**Scale Type**

Header: Hist:ScaleType  
Type: Selector  
Arguments: Linear, Logarithm  
Action: Sets the type of histogram scale

**Scale Mode**

Header: Hist:ScaleMode  
Type: Selector  
Arguments: Auto, Manual  
Action: Sets the mode of histogram scale

***Linear Scale of Vertical or Horizontal Histogram***

Headers: Hist:VertScale  
Hist:HorScale

Type: Float

Argument: (10 to 100) %/div

Action: Sets the scale of histograms. Used for Manual mode and Linear type of scale only.

***Linear Offset of Vertical or Horizontal Histogram***

Headers: Hist:VertOffset  
Hist:HorOffset

Type: Float

Argument: 0% to 100%

Action: Sets the offset of the histograms. It used for Manual mode and Linear type of scale only.

***Logarithmic Scale of Vertical or Horizontal Histogram***

Headers: Hist:VertDBScale  
Hist:HorDBScale

Type: Float

Argument: (6 to 60) dB/div

Action: Sets the scale of the histograms. Used for Manual mode and Logarithm type of scale only.

***Logarithmic Offset of Vertical or Horizontal Histogram***

Headers: Hist:VertDBOffs  
Hist:HorDBOffs

Type: Float

Argument: (-60 to 0) dB

Action: Sets the offset of the histograms. Used for Manual mode and Logarithm type of scale only.

#### 4.14.6 Histogram Result Commands

##### ***Get Histogram Data***

Headers: Hist:Data?

Type: Data

Argument: None

Forms: Query only

Action: Returns a set of text strings with the pair of numbers (comma-separated). First number in the each pair is the histogram axis value, and second number is the histogram value in this point.

##### ***Get Histogram Measure***

Headers: Hist:Res:<Param>?

Parameter<Param>:

- InBox – number of hints in box
- Wfm – number of waveforms
- Peak – peak value of histogram
- PP – difference between highest and lowest values of signal
- Median – centre between highest and lowest values of signal
- Mean – average of distribution of histogram
- StdDev – standard deviation of histogram
- Mean1S – number of hints in Mean  $\pm$  StdDev region, %
- Mean2S – number of hints in Mean  $\pm$  2StdDev region, %
- Mean3S – number of hints in Mean  $\pm$  3StdDev region, %
- Min – min. value of signal
- Max – max. value of signal
- Max-Max – difference between two values of signal, matched two max of histogram

Type: Data

Argument: None

Forms: Query only

Action: Returns a text string with the value of the specified parameter

## 4.15 Eye Diagram commands

### 4.15.1 General Eye Commands

#### **Type of Eye Measurements**

Header: Eye:Measure

Type: Selector

Arguments: Off, NRZ, RZ

Action: Sets the type of eye measurements

#### **Sources for Eye Measurements**

Header: Eye:Source

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Sets the source for eye measurements

#### **Number of Waveforms in one Measurement**

Header: Eye:WfmsInCycle

Type: Integer

Argument: 64, 128, 256, 512, 1024

Action: Sets the number of waveforms in one measurement

### 4.15.2 Eye Measurements Commands

#### **List of X-Axis NRZ Measurements**

Header: Eye:XNRZParam

Type: Group-on/off

Items: Area, BitRate, BitTime, CrossTime, CycleArea, DutCycDistP, DutCycDists, EyeWidth, EyeWidthP, FallTime, Freq, JitterPP, JitterRMS, Period, RiseTime

Action: Defines the set of X-axis measurements for NRZ signals

***List of Y-Axis NRZ Measurements***

Header: Eye:YNRZParam  
Type: Group-on/off  
Items: AcRMS, AvgPower, AvgPWDdBm, CrossPerc, CrossLevel, ExtRatioDB, ExtRatioP, ExtRatio, EyeAmpl, EyeHeight, EyeHeightDB, Max, Mean, Mid, Min, NegOver, PPNoiseOne, PPNoiseZero, RMSNoiseOne, RMSNoiseZero, OneLevel, PeakPeak, PosOver, RMS, SNRaio, SNRaioDB, ZeroLevel  
Action: Defines the set of Y-axis measurements for NRZ signals

***List of X-Axis RZ Measurements***

Header: Eye:XRZParam  
Type: Group-on/off  
Items: Area, BitRate, BitTime, CycleArea, EyeWidth, EyeWidthP, FallTime, JittPpFall, JittPpRise, JittRMSFall, JittRMSRise, NegCross, PosCross, PosDutyCyc, PulseSymm, PulseWidth, RiseTime  
Action: Defines the set of X-axis measurements for RZ signals

***List of Y-Axis RZ Measurements***

Header: Eye:YRZParam  
Type: Group-on/off  
Items: AcRMS, AvgPower, AvgPWDdBm, Contrast, ContrastBb, ContrastP, ExtRatioDB, ExtRatioP, ExtRatio, EyeAmpl, EyeHeight, EyeHeightDB, EyeOpenFact, Max, Mean, Mid, Min, PPNoiseOne, PPNoiseZero, RmsNoiseOne, RMSNoiseZero, OneLevel, PeakPeak, RMS, SignToNoise, ZeroLevel  
Action: Defines the set of Y-axis measurements for RZ signals

***Measurements List Clearing***

Header: Eye:ClearAllMeas  
Type: Executing  
Action: Clears the list of measurement parameters

#### 4.15.3 Define Parameters Commands

##### ***Eye Frame Visibility***

Header: Eye:DisplayWind

Type: On/off

Action: Sets the visibility of the eye frame

##### ***Left and Right Boundary for NRZ Top/Base Finding***

Headers: Eye:LeftBound

Eye:RightBound

Type: Float

Argument: 10% to 90% of the NRZ period

Action: Sets the zone of the period of the NRZ signal for the top/base calculation

##### ***Threshold Definition Mode***

Header: Eye:ThreshMode

Type: Selector

Arguments: 10-90, 20-80, Custom

Action: Sets the mode of threshold definition

##### ***Upper and Lower Threshold***

Headers: Eye:UpThresh

Eye:LowThresh

Type: Float

Argument: 5% to 95% of amplitude

Action: Sets the thresholds for the slopes calculation. Used for Custom mode.

#### 4.15.4 Eye Calculation Commands

##### ***Measurement Statistic***

Header: Eye:Statistic

Type: On/off

Action: Enables/disables measurement statistics

##### ***Measurement Statistic Mode***

Header: Eye:Mode

Type: Selector

Arguments: Permanent, Window, Weight

Action: Sets the mode of statistics calculation. Used when statistic is enable.

##### ***Windows Value***

Header: Eye:Window

Type: Integer

Argument: 8, 16, 32, ..., 8192

Action: Sets the window value. Used for Window mode of statistics.

##### ***Weight Value***

Header: Eye:Weight

Type: Integer

Argument: 8, 16, 32, ..., 8192

Action: Sets the weight value. Used for Weight mode of statistics.

#### 4.15.5 Getting Eye Measurement Results

##### ***Get List of Measured Parameters***

Header: Eye:Res>List?

Type: Data

Argument: None

Forms: Query only

Action: Returns a list of active eye measurements with ordinal index

##### ***Get Current Value of Parameter***

Header: Eye:Res:<N>?

Parameter <N>: Index of parameter in the list

Type: Data

Argument: None

Forms: Query only

Action: Returns the result of the specified measured parameter

##### ***Get Statistic Value of Parameter***

Header: Eye:Res:<N>:<Val>?

Parameter <N>: index of the parameter in the list

Parameter <Val>: TotalMeas, Min, Max (when Measurement Statistic off)  
Cycle, Wfm, Min, Max, Mean, StdDev (when Measurement Statistic on)

Type: Data-type command

Arguments: none

Forms: query only.

Action: return the specified statistical parameter of the measured parameter

## 4.16 Mask Test commands

### 4.16.1 Common Mask Test Commands

#### ***Mask Test On***

Header: Mask:TestOn

Type: On/off

Action: Enables/disables the mask test functionality

#### ***Signal for Mask Testing***

Header: Mask:CompareWith

Type: Selector

Arguments: Ch1, Ch2, CH3, CH4, F1, F2, F3, F4, DB

Action: Selects the signal for mask testing

#### ***ActuateMask Testing***

Header: Mask:Testing

Type: On/off

Action: Enables/disables the comparison with current mask

#### ***Mask Erasing***

Header: Mask:EraseMask

Type: Execution

Action: Clears the current mask from the display

### 4.16.2 Mask Creating

#### ***Mask Creating Mode***

Header: Mask>CreateAs

Type: Selector

Arguments: Std, Auto, Edit

Action: Sets the mask creation method

#### 4.16.3 Standard Mask Test Commands

##### ***Get List of Standards***

Header: StdMask:StdList?

Type: Data

Argument: None

Forms: Query only

Action: Returns a list of mask standards with ordinal index

##### ***Select Standard***

Header: StdMask:StdIndex

Type: Integer

Argument: 0 to (number of standards-1)

Action: Selects the current standard by its ordinal index

##### ***Get List of Masks***

Header: StdMask:MasksList?

Type: Data

Argument: None

Forms: Query only

Action: Returns a list of masks with ordinal index from the selected standard

##### ***Select Standard Mask***

Header: StdMask:MaskIndex

Type: Integer

Argument: 0 to (number of masks in the current standard-1)

Action: Loads the specified mask by its ordinal index

##### ***Alignment of Signal with Standard Mask***

Header: StdMask:Align

Type: On/off

Action: Enables/disables the alignment of the tested signal with the standard mask parameters

***Enable Margins***

Header: StdMask:MarginsOn

Type: On/off

Action: Enables/disables the margin control of eye-typed masks

***Margins Value***

Header: StdMask:MarginsVal

Type: Float

Arguments: -100% to +100%

Action: Sets the margin's value. Used when margins are enabled

***Build Immediately***

Header: StdMask:BuildImmediate

Type: On/off

Action: Enables/disables creation of the standard mask immediately after any of its parameters change

#### 4.16.4 Automask Commands

##### ***Automask Source***

Header: Automask:Source

Type: Selector

Arguments: Ch1, Ch2, CH3, CH4, F1, F2, F3, F4

Action: Selects the signal as a template for automask building

##### ***Margins Units***

Header: Automask:Unit

Type: Selector-type command

Arguments: Division, Current

Action: Selects the margins units for automask building

##### ***Automask X-Margins***

Header: Automask:DeltaX

Type: Float

Arguments: (0.02 to 2) div for Division margins units real X-axis value for Current margins units

Action: Sets the X-axis margins around the template signal

##### ***Automask Y-Margins***

Header: Automask:DeltaY

Type: Float

Arguments: (0.03125 to 2) div for Division margins units real Y-axis value for Current margins units

Action: Sets the Y-axis margins around the template signal

##### ***Automask Build***

Header: Automask:BuildAMask

Type: Execution

Action: Builds automask immediately

#### 4.16.5 Mask Test Termination

##### ***Mask Test Finish Condition***

Header: Mask:RunUntil

Type: Selector

Arguments: StopBtn, FailedWfms, FailedSmpls, Wfms, Samples

Action: Sets the condition of mask test termination

##### ***Number of Failed Waveforms***

Header: Mask:FailWfms

Type: Integer

Argument: 1 to 1000000

Action: Sets the number of failed waveforms for the FailedWfms finish condition

##### ***Number of Failed Samples***

Header: Mask:FailSmpls

Type: Integer-type command

Argument: 1 to 1000000

Action: Sets the number of failed samples for the FailedSmpls finish condition

##### ***Number of Waveforms***

Header: Mask:NWfms

Type: Integer-type command

Argument: 1 to 1000000

Action: Sets the number of waveforms for the Wfms finish condition

##### ***Number of Samples***

Header: Mask:NSamples

Type: Integer-type command

Argument: 1 to 1000000

Action: Sets the number of samples for the Samples finish condition

#### 4.16.6 Mask Test Actions

##### **Select Actions**

Header: Mask:Action

Type: Group-on/off

Items: Beep, Save

Action: Save – every failed signal is stored to disk  
Beep – the beep signal will sound for every failed signal

##### **Format of Stored Files**

Header: Mask:FileFormat

Type: Selector

Arguments: Binary, Verbose, YOnly

Action: Sets the file format. Used when Save action is on.

##### **Stored File Name**

Header: Mask:FileName

Type: Data

Argument: Text string

Forms: Command, query, command with query

Action: Defines the name for storing failed signals on Disk. Used when Save action is on.

#### 4.16.7 User Mask

##### **User Masks File Name**

Header: Mask:MaskFile

Type: Data

Argument: Text string

Forms: Command, query, command with query.

Action: Defines the file name for next loading or saving user mask from the disk

##### **Load User Mask**

Header: Mask:LoadUser

Type: Execution

Action: Loads the previously specified user mask

***Save User Mask***

Header: Mask:SaveUser

Type: Execution

Action: Saves the current mask as user with previously specified file name

## 4.16.8 Getting Mask Test Results

***Get Integrated Results of Mask Test***

Headers: Mask:Res:<Param>?

Parameter <Param>:

AllWfm	- number of waveforms
FailWfm	- number of failed waveforms
AllSmpl	- number of samples
FailSmpl	- number of failed samples

Type: Data

Argument: None

Forms: Query only

Action: Returns a text string with the value of the specified parameter

***Get Number of Samples in Selected Polygons***

Headers: Mask:Res:Poly<N>?

Parameter <N>: Number of the polygon, 1 to 8

Type: Data

Argument: None

Forms: Query only

Action: Returns a text string with the values of failed samples on specified polygon

***Get Number of Samples in Margins of Selected Polygon***

Headers: Mask:Res:Poly<N>Mar?

Parameter <N>: Number of the polygon, 1 to 4

Type: Data

Argument: None

Forms: Query only

Action: Returns a text string with the values of failed samples on the margin of specified polygon. Used when Margins enabled.

***Get Number of Samples in Selected Polygon with Margins Together***

Headers: Mask:Res:Poly<N>All?

Parameter <N>: Number of the polygon, 1 to 4

Type: Data

Argument: None

Forms: Query only

Action: Returns a text string with the total number of failed samples on the margin and on the specified polygon. Used when Margins enabled.

## 4.17 Autocalibration commands

### 4.17.1 Single-shot Autocalibration

#### ***Start autocalibration of channels 1 and 2***

Header: Flash:Sampler:Autocal:Start

Type: Executing-type command

Action:Start self-calibration of Sampler 1

#### ***Start autocalibration of channels 3 and 4 (PicoScope 9341 only)***

Header: Flash:Smpllr2:Autocal:Start

Type: Executing-type command

Action:Start self-calibration of Sampler 2

#### ***Start autocalibration of timebase***

Header: FLASH:TBTrig:Autocal

Type: Execution

Action: Starts self-calibration of timebase

#### ***Get the autocalibration status query***

Header: Flash:Calibr:AutocalResult?

Type: Integer

Action: Command is ignored, and query returns an integer:

- 1 – Autocalibration in progress.
- 0 – Autocalibration finished OK.
- 1 – Signal must be disconnected from Ch1 Input. Autocalibration of the Channels is aborted.
- 2 – Signal must be disconnected from Ch2 Input. Autocalibration of the Channels is aborted.
- 3 – Signal must be disconnected from Ch1 and Ch2 Inputs. Autocalibration of the Channels is aborted.
- 5 – Autocalibration failed.

#### 4.17.2 Periodic Autocalibration

##### ***When to Begin Autocalibration***

Header: Util:CalibrWhen

Type: On/off-group

Items: PowerOn, Period, Temperat

Action: PowerOn – autocalibration begins on the next Power On  
 Period – autocalibration begins periodically after the specified interval  
 Temperat – autocalibration begins when deviation of temperature inside the instrument exceeds the specified value

Note. Periodic autocalibration must be turned off when GUI is in RemoteOnly or Invisible state. See [GUI command](#).

##### ***Autocalibration Period***

Header: Util:CalPeriod

Type: Float

Argument: 0.5 to 16 hours

Action: Sets the autocalibration period in hours

##### ***Temperature Deviation***

Header: Util:TempChange

Type: Float

Argument: 0.5 to 10 °C

Action: Sets the temperature deviation for autocalibration

##### ***Get the Temperature of the Instrument Query***

Header: Calibr:Temperature?

Type: Float

Argument: None

Forms: Query only

Action: Returns the temperature inside the device in degrees Celsius

#### 4.17.3 Balancing the channels manually

##### ***Balancing channels 1 and 2 manually***

Header: Flash:Sampler:Ch1:FullBW:Balance  
Flash:Sampler:Ch1:NarrowBW:Balance  
Flash:Sampler:Ch2:FullBW:Balance  
Flash:Sampler:Ch2:NarrowBW:Balance

Type: Float

Arguments: -0.5 to 0.5

Action: Query or set the balance value in volts for the specified channel

##### ***Balancing channels 3 and 4 manually (PicoScope 9341 only)***

Header: Flash:Smpllr2:Ch3:FullBW:Balance  
Flash:Smpllr2:Ch3:NarrowBW:Balance  
Flash:Smpllr2:Ch4:FullBW:Balance  
Flash:Smpllr2:Ch4:NarrowBW:Balance

Type: Float

Arguments: -0.5 to 0.5

Action: Query or set the balance value in volts for the specified channel of Sampler 2

## 4.18 Waveforms commands

This group of commands is designed for receiving acquired waveforms from the oscilloscope.

### **Waveform Source**

Header: Wfm:Source

Type: Selector

Arguments: Ch1, Ch2, Ch3, Ch4, F1, F2, F3, F4, M1, M2, M3, M4

Action: Sets the signal to be received

### **Spectrum Format**

Header: Wfm:Complex

Type: Selector

Arguments: Mod, Ph, Re, Im

Action: Selects the received component of the complex signal. Used for spectrum data.

### **Get Waveform Data**

Header: Wfm:Data?

Type: Data

Argument: None

Forms: Query only

Action: Returns a text string with values of all points of the signal (comma-separated)

### **Get Number of Points in the Waveform**

Header: Wfm:Preamb:Poin?

Type: Data

Argument: None

Forms: Query only

Action: Returns the number of points in the signal

***Get X-Axis Step***

Header: Wfm:Preamb:XInc?

Type: Data

Argument: None

Forms: Query only

Action: Returns the increment on the X-axis for one signal point

***Get X-Axis Origin***

Header: Wfm:Preamb:XOrg?

Type: Data

Argument: None

Forms: Query only

Action: Returns the X-axis value for the first signal point

***Get X-Axis Unit***

Header: Wfm:Preamb:XU?

Type: Data

Argument: None

Forms: Query only

Action: Returns the X-axis physical units

***Get Y-Axis Unit***

Header: Wfm:Preamb:YU?

Type: Data

Argument: None

Forms: Query only

Action: Returns the Y-axis physical units

## 4.19 Optical commands (PicoScope 9321 only)

### 4.19.1 Common optical commands

#### **O/E converter commutation**

Header: OEConv:Dest

Type: Selector

Arguments: OEConv, Ch1, Ch2

Action: set destination of the O/E converters stimuli

#### **Current wavelength**

Header: OEConv:Wavelen

Type: Selector

Arguments: WL1, WL2, WL3, WLUserDef

Action: set the current wavelength.

Note: Arguments WL1, WL2, WL3 selects one of the factory defined wavelength value. Usually WL1 is 1550 nm, WL2 is 1310 nm and WL3 is 850 nm. Argument WLUserDef allows you to select user defined values.

#### **Select user wavelength value**

Header: OEConv:UserWaveLen

Type: Selector

Arguments: WL1, WL2, WL3, ..., WL8

Action: set the current user defined wavelength.

Note: All user-defined wavelengths are sorted by value. WL1 is the shortest user-defined wavelength.

#### **Reading current O/E conversion gain value**

Header: OEConv:ConversionGain

Type: Float

Argument: 10 V/W to 1500 V/W

Forms: query only

Action: reading the current conversion gain value.

#### 4.19.2 User-defined calibration points commands

##### ***Erasing all user-defined calibration points***

Header: Flash:OeConv:ClearAllPoints  
Type: Execution  
Action: Erase all user-defined wavelength points.

##### ***Erasing one user-defined calibration point***

Header: Flash:OeConv:ClearPoint1  
Flash:OeConv:ClearPoint2  
...  
Flash:OeConv:ClearPoint8  
  
Type: Execution  
  
Action: Erase specified user-defined wavelength point.

##### ***Set the wavelength value for user-defined calibration point***

Header: Flash:OeConv:CalWavelen1  
Flash:OeConv:CalWavelen2  
...  
Flash:OeConv:CalWavelen8  
  
Type: Float  
  
Argument: 750 nm to 1.65 μm  
  
Action: setting the wavelength value for user-defined point.

##### ***Set the conversion gain value for user-defined calibration point***

Header: Flash:OeConv:CalGain1  
Flash:OeConv:CalGain2  
...  
Flash:OeConv:CalGain8  
  
Type: Float  
  
Argument: 10 V/W to 1500 V/W  
  
Action: setting the conversion gain value for user-defined point.

#### 4.19.3 Dark Level calibration commands

##### ***Dark Level autocalibration***

Header: Flash:OeConv:AutoDarkLvl  
Type: Execution  
Action: starting the Dark Level autocalibration procedure. The duration of this process is less than 5 s.

***Dark Level value***

Header: Flash:OeConv:DarkLvl

Type: Float

Argument: 0 % to 100 %

Action: query about Dark Level value after autocalibration procedure; setting Dark Level value when autocalibration is undesirable.

#### 4.19.4 Calibration values saving commands

***Store all calibration values***

Header: Flash:OeConv:Calibr:Store

Type: Execution

Action: Store current Dark Level and all user defined calibration points into the nonvolatile memory of the PicoScope 9321.

***Recall all calibration values***

Header: Flash:OeConv:Calibr:Recall

Type: Execution

Action: Recall current Dark Level and all user defined calibration points from the nonvolatile memory of the PicoScope 9321. This command calls automatically when PicoScope 9321 power on.

## 4.20 TDR/TDT Commands

### 4.20.1 Stimuli source command (PicoScope 9311 and 9312 only)

#### ***Stimuli source command***

Header: TDR:Stimul:Source

Type: Selector

Arguments: Internal, External

Action: set source of stimuli

### 4.20.2 Common Internal Stimuli commands (PicoScope 9311/9312 only)

#### ***Stimuli Mode***

Header: TDR:Stimul:G1VsG2

Type: Selector

Arguments: Separate, Paired

Action: set mode of stimulus setups

#### ***Internal Stimuli On***

Header: TDR:Stimul:Out1:On                   TDR:Stimul:Out2:On

Type: On/off

Action: turn on or turn off the specified stimulus

#### ***Internal Stimuli Period***

Header: TDR:Stimul:Period

Type: Float

Argument: 1e-6 to 60e-3

Action: set pulse period of the stimuli in s

#### ***Internal Stimuli Width***

Header: TDR:Stimul:Width

Type: Float

Argument: 2e-7 to 4e-6

Action: set pulse width of the stimuli in s

#### 4.20.3 PicoScope 9311 Internal Stimuli commands

##### ***Internal Stimuli Amplitude***

Header: TDR:Stimul:Srd:Out1:Ampl  
TDR:Stimul:Srd:Out2:Ampl

Type: Float-type command

Argument: 2 to (8); max value depends on the adjustments of generators.

Action: set stimuli amplitude in V

##### ***Internal Stimuli Deskew***

Header: TDR:Stimul:Deskew

Type: Float

Argument: -500e-12 to 500e-12

Action: set stimuli deskew in s

##### ***Internal Stimuli pre-trigger***

Header: TDR:Stimul:Pretrig

Type: Float

Argument: 20e-9 to 25e-9

Action: set stimuli pre-trigger in s

##### ***Internal Stimuli performance***

Header: TDR:Stimul:Srd:Out1:Perform  
TDR:Stimul:Srd:Out2:Perform

Type: Selector

Arguments: Flat, Fast

Action: set performance of the specified stimulus

#### 4.20.4 PicoScope 9312 Internal Stimuli commands

##### ***Internal Stimuli Deskew***

Header: TDR:Stimul:Deskew

Type: Float

Argument: 0 % to 100 %

Action: set stimuli deskew in %

##### ***Tunnel diode head calibration***

Header: TDR:Stimul:Td:Out1:Autocal  
TDR:Stimul:Td:Out2:Autocal

Type: Execution

Action: find the optimal tunnel diode current

#### **Tunnel diode current mode**

Header: TDR:Stimul:Td:Out1:CurrentMode  
TDR:Stimul:Td:Out2:CurrentMode

Type: Selector

Arguments: Auto, Manual

Action: set mode of tunnel diode current control for the specified stimulus

#### **Tunnel diode current**

Header: TDR:Stimul:Td:Out1:Sensitive  
TDR:Stimul:Td:Out2:Sensitive

Type: Float

Argument: 0 to 0.06 for positive head; -0.06 to 0 for negative head

Action: set the tunnel diode current in A

### 4.20.5 Reflectometer Setup commands

#### **Reflectometer Mode**

Header: TDR:Mode

Type: Selector

Arguments: Off, TDR, TDT

Action: set reflectometer mode

#### **Device under Test**

Header: TDR:DUT

Type: Selector

Arguments: Single, Differ, Common

Action: set the device under test.

Note: both the stimuli must be on for *Differ* and *Common* modes

#### **Stimulus Destination**

Header: TDR:Ch1:Destination  
TDR:Ch2:Destination

Type: Selector

Arguments: G1, G2, Off

Action: set the stimulus connected to the specified channel.

### **Polarity of the stimulus**

Header: TDR:Ch1:Polar  
TDR:Ch2:Polar

Type: Selector

Arguments: Pos, Neg

Action: get the polarity of the internal stimulus of the specified channel; set polarity of the external stimulus.

### **Vertical scale units for TDR mode**

Header: TDR:Ch1:VertScaleTDR  
TDR:Ch2:VertScaleTDR

Type: Selector

Arguments: Volt, Rho, Ohm

Action: set vertical scale for TDR mode.

### **Vertical scale units for TDT mode**

Header: TDR:Ch1:VertScaleTDT  
TDR:Ch2:VertScaleTDT

Type: Selector

Arguments: Volt, Gain, DbGain

Action: set vertical scale for TDT mode.

### **Horizontal Scale units**

Header: TDR:HorScale

Type: Selector

Arguments: Time, Meter, Foot, Inch

Action: set horizontal scale units.

### **Mode of transmission line parameters**

Header: TDR:Unit

Type: Selector

Arguments: Velocity, DielConst

Action: set mode of transmission line parameter definitions. It used for non-time horizontal scale units

***Propagation Velocity***

Header: TDR:Veloc  
Type: Float  
Argument: 0.1 to 1  
Action: set propagation velocity value for the transmission line.

***Dielectric Constant***

Header: TDR:DielConst  
Type: Float  
Argument: 1 to 100  
Action: set Dielectric Constant value for the transmission line.

***Reference Amplitude (TDR Mode)***

Header: TDR:Ch1:RefAmplTDR  
          TDR:Ch2:RefAmplTDR  
  
Type: Float  
  
Argument: 0.001 V to 10 V  
  
Action: set Reference Amplitude value for calculating Rho, Ohm.

***Base Line (TDR Mode)***

Header: TDR:Ch1:BaseLineTDR  
          TDR:Ch2:BaseLineTDR  
  
Type: Float  
  
Argument: -10 V to 10 V  
  
Action: set Base Line value for calculating Rho, Ohm.

***Reference Amplitude (TDT Mode)***

Header: TDR:Ch1:RefAmplTDT  
          TDR:Ch2:RefAmplTDT  
  
Type: Float  
  
Argument: -10 V to 10 V  
  
Action: set Reference Amplitude value for calculating Gain and Gain (dB).

***Reference Zero (TDT Mode)***

Header: TDR:Ch1:RefZeroTDT  
          TDR:Ch2:RefZeroTDT  
  
Type: Float

Argument: -10 V to 10 V

Action: set Reference Zero value for calculating Gain and Gain (dB).

### **Reference Plane**

Header: TDR:Ch1:RefPlane  
TDR:Ch2:RefPlane

Type: Float

Argument: -1e-11 s to 1e-8 s

Action: set Reference plane position.

### **Corrected Time**

Header: TDR:Ch1:CorrectTime  
TDR:Ch2:CorrectTime

Type: Float

Argument: -1e-11 s to 1e-8 s

Action: set the Corrected Time for pulse correction.

### **Start Calibration Rho (TDR Mode)**

Header: TDR:Ch1:DoCalibrTDR  
TDR:Ch2:DoCalibrTDR

Type: Execution

Action: Rho calibration procedure. Follow the instructions on the screen.

Note before sending this command, check:

- Reflectometer Mode is TDR;
- specified channel is active;
- selected stimulus is ON.

### **Start Gain Calibration (TDT Mode)**

Header: TDR:Ch1:DoCalibrTDT  
TDR:Ch2:DoCalibrTDT

Type: Execution

Action: Gain Calibration procedure. Follow the instructions on the screen.

Note before sending this command, check:

- Reflectometer Mode is TDT;
- specified channel is active;
- selected stimulus is ON.

### **Correction**

Header: TDR:Ch1:Correction  
TDR:Ch2:Correction

Type: On/Off

Action: switch on Correction mode. Do this immediately after switching on. Follow the instructions on the screen.

Note before sending this command, check:  
- Reflectometer Mode is TDR or TDT;  
- specified channel is active;  
- selected stimulus is ON.

## 4.21 Instrument Info Commands

### ***Get Device Model***

Header: GetInfo:Model?

Type: Data

Argument: none

Forms: query only

Action: return text string with model of the instrument

### ***Get Device Functionality***

Header: GetInfo:Function?

Type: Data

Argument: none

Forms: query only

Action: return text string with functionality of the instrument

### ***Get Device Year***

Header: GetInfo:Year?

Type: Data

Argument: none

Forms: query only

Action: return text string with year of the instrument

### ***Get Device Serial Number***

Header: GetInfo:SerialNr?

Type: Data

Argument: none

Forms: query only

Action: return text string with Serial Number of the instrument

***Get Device FPGA FW Version***

Header: GetInfo:HwVersion?

Type: Data

Argument: none

Forms: query only

Action: return text string with FPGA firmware version of the instrument

***Get Device FW Version***

Header: GetInfo:FwVersion?

Type: Data

Argument: none

Forms: query only

Action: return text string with firmware version of the instrument

***Get Device SW Version***

Header: GetInfo:SwVersion?

Type: Data

Argument: none

Forms: query only

Action: return text string with software version of the instrument

***Get Current Interface of Device***

Header: GetInfo:Interface?

Type: Data

Argument: none

Forms: query only

Action: return text string with current interface of the instrument

***Get Current IP Address***

Header: GetInfo:IpAddress?

Type: Data

Argument: none

Forms: query only

Action: return text string with the current IP address of the instrument

***Get Current Subnet Mask***

Header: GetInfo:SubnetMask?

Type: Data

Argument: none

Forms: query only

Action: return text string with the current subnet mask of the instrument

***Get Current Gateway***

Header: GetInfo:Gateway?

Type: Data

Argument: none

Forms: query only

Action: return text string with current gateway of the instrument

***Get Current MAC Address***

Header: GetInfo:MacAddress?

Type: Data

Argument: none

Forms: query only

Action: return text string with the current MAC address of the instrument

## 5 Revision history

Date	Version	Changes
2017-02-09	4	<p><b>4.2 GUI commands</b> Chapter fully updated</p> <p><b>4.4 Channels commands</b></p> <ul style="list-style-type: none"> <li>a. Removed space from AllChs : BestFlat mnemonics</li> <li>b. Added <b>Scale a TDR Channel</b> command</li> <li>c. Added <b>Scale a TDT Channel</b> command</li> <li>d. Added <b>Offset a TDR Channel</b> command</li> <li>e. Added <b>Offset a TDT Channel</b> command</li> </ul> <p><b>4.5 Timebase commands</b></p> <ul style="list-style-type: none"> <li>a. Added <b>Main Distances Scale for TDR-TDT</b> command</li> <li>b. Added <b>Delayed Distances Scale for TDR-TDT</b> command</li> <li>c. Added <b>Distances Delay for TDR-TDT</b> command</li> </ul> <p><b>4.8 Display commands</b></p> <ul style="list-style-type: none"> <li>a. Added <b>Display Persistence (SW v.3.20 or newer)</b> command</li> <li>b. Added <b>Display Style (SW v.3.20 or newer)</b> command</li> <li>c. Updated arguments of <b>Graticule Type</b> command</li> </ul> <p><b>4.9.3 Work with Disk</b> Updated header of <b>Save to Disk</b> command</p> <p><b>4.10 Markers commands</b> Added <b>Y value of XY-Marker</b> query</p> <p><b>4.11 Measure commands</b> Structure of Chapter updated</p> <p><b>4.11.2 Statistic Commands</b> Added <b>Clear Statistics for all Measures</b> command</p> <p><b>4.13.5 Function Operands</b> Updated arguments of <b>Operand 2</b> command</p> <p><b>4.15.5 Getting Eye Measurement Results</b> Updated <b>Get Statistic Value of Parameter</b> command</p> <p><b>4.16.1 Common Mask Test Commands</b> Removed space from Mask : Testing mnemonics</p> <p><b>4.16.4 Automask Commands</b> Removed space from Automask : BuildAMask mnemonics</p> <p><b>4.19.2 User-defined calibration points commands</b></p> <ul style="list-style-type: none"> <li>a. Updated header of <b>Set the wavelength value for user-defined calibration point</b> command</li> <li>b. Updated header of <b>Set the conversion gain value for user-defined calibration point</b> command</li> </ul> <p><b>4.20.2 Common Internal Stimuli commands</b> Added <b>Internal Stimuli On</b> command</p> <p><b>4.20.3 PicoScope 9311 Internal Stimuli commands</b></p> <ul style="list-style-type: none"> <li>a. Added <b>Internal Stimuli Amplitude</b> command</li> <li>b. Updated header of <b>Internal Stimuli performance</b> command</li> </ul>

		<p><b>4.20.5 Reflectometer Setup commands</b></p> <ul style="list-style-type: none"><li>a. Added <b>Start Calibration Rho</b> command</li><li>b. Added <b>Start Gain Calibration</b> command</li><li>c. Added <b>Correction</b> command</li></ul>
		<p><b>4.21 Instrument Info Commands</b></p> <p>New Chapter added</p>
		<p><b>5 Revision history</b></p> <p>New Chapter added</p>



## Index

### A

Acquire a Channel  
     Attenuator dB 23  
     Attenuator linear/log 23  
     Attenuator ratio 23  
     Attenuator unit 23  
 Acquisition commands  
     Acquisition Mode of Channel 25  
     Action when Number of Waveforms reached 25  
     Channel Averaging 25  
     Channel Envelopes 25  
     Channel Record Length 25  
     Number of Waveforms 25  
     Sampling Mode 25  
     Termination of Acquisition 25  
     Type of signal 25  
 All-off mode 6

### C

Case-insensitivity 4  
 Channels commands  
     Acquire a Channel 13  
     Attenuator dB 13  
     Attenuator linear/log 13  
     Attenuator ratio 13  
     Attenuator unit 13  
     Bandwidth of Channel 13  
     Channels commands 13  
     Display a Channel 13  
     Offset a Channel 13  
     Scale a Channel 13

Characteristics 6  
 Command classification 6  
 Command entry rules 4  
 Command header 3  
 Command messages 2  
 Commands 3  
 Common Mask Commands  
     Actuate Mask Testing 69  
     Format of Stored Files 69  
     Load User Mask 69  
     Mask Erasing 69  
     Mask Test Actions 69  
     Mask Test Finish Condition 69  
     Number of Failed Samples 69  
     Number of Failed Waveforms 69

Number of Samples 69  
 Number of Waveforms 69  
 Signal for Mask Testing 69  
 Stored File Name 69  
 User Masks File Name 69  
 COMRC Object 1, 1  
 Concatenation 4

**D**  
 Data-type commands 9  
 Display commands  
     Define Trace Screen (for 2YT, Comb2YTXY Formats) 28  
     Define Trace Screen (for 4YT Format) 28  
     Display commands 28  
     Display Format 28  
     Graticule Type 28  
     Persistence Time, seconds (for VarPersist Style) 28  
     Refresh Time, seconds (for VarGrayScal or VColorGrade Styles) 28  
     Reset Display Style 28  
     Select active trace 28  
     Set Display Style 28  
     Source of X Axis for XY Screen 28  
     Source of Y Axis for XY Screen 28  
     Trace mode 28

### E

ExecCommand Method 1  
 Execution-type commands 6

### F

Float-type commands 9

### G

Getting Eye Measurement Results, commands for  
     Get Current Value of Parameter 68  
     Get List of Measured Parameters 68  
     Get Statistic Value of Parameter 68  
 Getting Mask Results, commands for  
     Get Integrated Results of Mask Test 75  
     Get Number of Samples in Margins of Selected Polygon 75  
     Get Number of Samples in Selected Polygon with Margins Together 75  
     Get Number of Samples in Selected Polygons 75  
 Getting Measurement Results, commands for  
     Get Current Value of Parameter 45  
     Get List of Measured Parameters 45

Getting Measurement Results, commands for  
Get Statistic Value of Parameter 45

Group-off mode 6

Group-on mode 6

Group-query mode 6

GUI commands 10

## H

Header commands  
Header 10

Headers 3

## I

Integer-type commands 8

## L

Limit Tests commands  
Action 46  
Action If 46  
Center Value 46  
Failure When 46  
File Name 46  
Format of Stored Files 46  
If Measurement Undefined 46  
Limit Test On/Off 46  
Limit Test Termination Condition 46  
Number of Failures 46  
Number of Waveforms 46  
Parameter Activity 46  
Parameter Center Mode 46  
Parameter Delta Mode 46  
Parameter Delta Percentage for User Defined mode 46  
Parameter Delta Value for Standard Deviation mode 46  
Parameter Delta Value for User Defined Mode 46  
Parameter Limit Mode 46  
Upper and Lower Limits of Parameters 46

Line feeds 4

## M

Markers commands  
Marker Sources 37  
Marker Type 37  
Motion of Markers 37  
X position of Marker 37  
Y position of Marker 37

Mathematics commands  
Constant Value 50

Enable Mathematical Function 50  
Function Operator 50  
Operand 1 50  
Operand 2 50  
Smoothing Parameter 50  
Trend Measurement 50

Measure commands  
Base Value (for Marker Method) 38  
Delete all Measures for all Sources 38  
Execute Single Measurement 38  
List of X Inter-Signal Measurements 38  
List of X Measurements 38  
List of Y Inter-Signal Measurements 38  
List of Y Measurements 38  
Margins Definition Mode 38  
Measurement Mode 38  
Measurement Source 38  
Measurement Type 38  
Measurements of Time Domain Signals commands 38  
Percentage of Upper, Middle or Lower Threshold 38  
Position of Left or Right Margin 38  
Position of Upper, Middle or Lower Threshold 38  
Second Source for Inter-Signal Measurements 38  
Slope of Left or Right Margins 38  
Statistic Measurement Mode 38  
Threshold Definition Method 38  
Threshold Units 38  
Thresholds of Left and Right Margin Slopes 38  
Top Value for Marker Method 38  
Top/Base Definition Method 38  
Viewing of Define Parameters 38  
Weight Value 38  
Windows Value 38

Measurement of Spectrum Signals, commands for  
Left and Right Spectrum Peaks 43  
Left and Right Spectrums Margin 43  
Limits Definition Method for Spectrum 43  
List of Spectrum Frequency Measurements 43  
List of Spectrum Magnitude Measurements 43  
Peak Level of Spectrum 43

Messages 2  
Mnemonics 4

## O

On/Off Group-type commands 6  
On/Off-type commands 6  
Overview 2

**P**

PicoScope9000 COM Server 1

**Q**

Queries 3

Query commands 2

**S**

Selector-type commands 8

Semicolon 4

Set commands 2

Single-item mode 6

Specifications 6

**Standard Mask Commands**

Alignment of Signal with Standard Mask 70

Enable Margins 70

Get List of Masks 70

Get List of Standards 70

Margins Value 70

Select Standard 70

Select Standard Mask 70

**System commands**

Clear Display 11

Recall Default Setup 11

Start Autoscaling 11

Start Cycle Acquisition 11

Start Single Acquisition / Stop Acquisition 11

System commands 11

**T**

**Timebase commands**

Acquire a Channel 17

Attenuator dB 17

Attenuator linear/log 17

Attenuator ratio 17

Attenuator unit 17

Bandwidth of Channel 17

Display a Channel 17

Offset a Channel 17

Scale a Channel 17

Timebase Commands 17

**Trigger Commands**

Bandwidth of Channel 20

Display a Channel 20

Offset a Channel 20

Scale a Channel 20

**U**

**Utilities commands**

Autocalibration Period 77

Start Autocalibration of Channels 77

Start Autocalibration of Timebase 77

Temperature Deviation 77

When to Begin Autocalibration 77

**W**

**Waveforms commands**

Get Number of Points in Waveform 80

Get Waveform Data 80

Get X-axis Origin 80

Get X-axis Step 80

Get X-axis Unit 80

Get Y-axis Unit 80

Spectrum Format 80

Waveform Source 80

White space 4

**Work with Disk commands**

File Name 33

File Name Mode 33

Format of stored files 33

Load 33

Save to Disk 33

Select Memory for loading signal from disk 33

Source for saving to file 33

**Work with Memo Zones (M1, M2, M3, M4) commands**

Memory Display 31

Save into Memory 31

Select Memory for Saving 31

Source for storing into Memory 31

**Work with Setups commands**

Name of Custom Setup File 35

Recall Custom Setup 35

Recall Factory Setup 35

Recall Power-Off Setup 35

Save Custom Setup 35

Save Setup as Default 35

**UK headquarters:**

Pico Technology  
James House  
Colmworth Business Park  
St. Neots  
Cambridgeshire  
PE19 8YP  
United Kingdom

Tel: +44 (0) 1480 396 395  
Fax: +44 (0) 1480 396 296

[sales@picotech.com](mailto:sales@picotech.com)  
[support@picotech.com](mailto:support@picotech.com)

[www.picotech.com](http://www.picotech.com)

**USA headquarters:**

Pico Technology  
320 N Glenwood Blvd  
Tyler  
Texas 75702  
United States

Tel: +1 800 591 2796  
Fax: +1 620 272 0981